

# Rajasthan Medical Services Corporation Limited

Gandhi Block, SwasthyaBhawan, Tilak Marg, C-Scheme, Jaipur-302005 (Raj.)

Phone No: 0141-2228066 ,2228064

Website: <http://rmsc.health.rajasthan.gov.in>

CIN:U24232RJ2011SGC035067

E\_mail :[edprmsc@gmail.com](mailto:edprmsc@gmail.com)

Ref. No.:- F.02(75)/RMSCL/PROC./N-95 Mask/NIB-07/2020/99

Dated:- 16/06/2020

## Corrigendum-II

**Subject: - Amended technical specification in bid and in BoQ.**

**Ref.:- NIB No. F.02(75)/RMSCL/PROC./N-95 Mask/NIB-07/2020/82 Dated:- 05.06.2020**


**(Technical bid opening due on dated -22.06.2020)**

The following amendments are made:

S. No	Existing specification	Amended specification												
9.	<b>ANNEXURE-1: BILL OF MATERIAL AND TECHNICAL SPECIFICATIONS and in BoQ</b>	<b>ANNEXURE-1: BILL OF MATERIAL AND TECHNICAL SPECIFICATIONS and in BoQ</b>												
	<table border="1"> <thead> <tr> <th>S.No.</th> <th>Description of Item and Technical specification</th> <th>Estimated Bid Qty.</th> </tr> </thead> <tbody> <tr> <td>1.</td> <td>                     N-95 Mask                     <ul style="list-style-type: none"> <li>Shape that will not collapse easily</li> <li>High filtration efficiency</li> <li>Good breathability, with expiratory valve</li> <li>Quality complaint with standards for medical N95 respirator:</li> <li>b. NIOSHN95, EN 149 FFP2</li> <li>Fluid resistance: minimum 80mmHg pressure based on ASTM F1862, ISO 22609</li> <li>Quality complaint with standards for particulate respirator that can be worn with full-face shield.</li> </ul> </td> <td>2500000</td> </tr> </tbody> </table>	S.No.	Description of Item and Technical specification	Estimated Bid Qty.	1.	N-95 Mask <ul style="list-style-type: none"> <li>Shape that will not collapse easily</li> <li>High filtration efficiency</li> <li>Good breathability, with expiratory valve</li> <li>Quality complaint with standards for medical N95 respirator:</li> <li>b. NIOSHN95, EN 149 FFP2</li> <li>Fluid resistance: minimum 80mmHg pressure based on ASTM F1862, ISO 22609</li> <li>Quality complaint with standards for particulate respirator that can be worn with full-face shield.</li> </ul>	2500000	<table border="1"> <thead> <tr> <th>S.No.</th> <th>Description of Item and Technical specification</th> <th>Estimated Bid Qty.</th> </tr> </thead> <tbody> <tr> <td>1.</td> <td>                     N-95 Mask                     <ul style="list-style-type: none"> <li>Shape that will not collapse easily</li> <li>High filtration efficiency</li> <li>Good breathability, with expiratory valve</li> <li>Quality complaint with standards for medical N95 respirator:</li> <li>b. NIOSHN95, EN 149 FFP2, or equivalent</li> <li>Fluid resistance: minimum 80mmHg pressure based on ASTM F1862, ISO 22609, or equivalent</li> <li>Quality complaint with standards for particulate respirator that can be worn with full-face shield.</li> </ul> </td> <td>2500000</td> </tr> </tbody> </table>	S.No.	Description of Item and Technical specification	Estimated Bid Qty.	1.	N-95 Mask <ul style="list-style-type: none"> <li>Shape that will not collapse easily</li> <li>High filtration efficiency</li> <li>Good breathability, with expiratory valve</li> <li>Quality complaint with standards for medical N95 respirator:</li> <li>b. NIOSHN95, EN 149 FFP2, or equivalent</li> <li>Fluid resistance: minimum 80mmHg pressure based on ASTM F1862, ISO 22609, or equivalent</li> <li>Quality complaint with standards for particulate respirator that can be worn with full-face shield.</li> </ul>	2500000
S.No.	Description of Item and Technical specification	Estimated Bid Qty.												
1.	N-95 Mask <ul style="list-style-type: none"> <li>Shape that will not collapse easily</li> <li>High filtration efficiency</li> <li>Good breathability, with expiratory valve</li> <li>Quality complaint with standards for medical N95 respirator:</li> <li>b. NIOSHN95, EN 149 FFP2</li> <li>Fluid resistance: minimum 80mmHg pressure based on ASTM F1862, ISO 22609</li> <li>Quality complaint with standards for particulate respirator that can be worn with full-face shield.</li> </ul>	2500000												
S.No.	Description of Item and Technical specification	Estimated Bid Qty.												
1.	N-95 Mask <ul style="list-style-type: none"> <li>Shape that will not collapse easily</li> <li>High filtration efficiency</li> <li>Good breathability, with expiratory valve</li> <li>Quality complaint with standards for medical N95 respirator:</li> <li>b. NIOSHN95, EN 149 FFP2, or equivalent</li> <li>Fluid resistance: minimum 80mmHg pressure based on ASTM F1862, ISO 22609, or equivalent</li> <li>Quality complaint with standards for particulate respirator that can be worn with full-face shield.</li> </ul>	2500000												

**Note:-**

- It may be noted that if any further amendments are issued then a corrigendum will be published and informed.
- Rest of the terms and conditions will remain the same.

  
**Executive Director (Proc.)**  
**RMSCL**  
