

Rajasthan Medical Services Corporation Limited

Gandhi Block, SwasthyaBhawan, Tilak Marg, C-Scheme, Jaipur-302005 (Raj.)

Phone No: 0141-2228066 ,2228064

Website: <http://rmsc.health.rajasthan.gov.in>

CIN:U24232RJ2011SGC035067

E_mail :edprmsc@nic.in

Ref. No.:- F.02(86)/RMSCL/Proc/N-95 mask /NIB-19/2020/469

Dated:- 18/12/2020

Corrigendum-III

Subject: - Amendment in specification.

Ref.:- NIB No. F.02(86)/RMSCL/Proc/N-95 mask /NIB-19/2020/442 Dated:01.12.2020

(Technical bid opening due on dated -29.12.2020)

The following amendments are made:

S. No	Existing Tender specification	Amended tender specification
1.	<p>ANNEXURE – VI and in BoQ</p> <p>N-95 Mask</p> <ul style="list-style-type: none"> • Shape that will not collapse easily • High filtration efficiency • Good breathability • without valve • Quality complaint with standards for medical N95 respirator: <ul style="list-style-type: none"> a. NIOSHN95,EN 149, FFP2/as per BIS 9473:2002 or equivalent • Fluid resistance: minimum 80mmHg pressure based on ASTM F1862, ISO 22609, or equivalent • Quality complaint with standards for particulate respirator that can be worn with full-face shield. 	<p>ANNEXURE – VI and in BoQ</p> <p>N-95 Mask</p> <ul style="list-style-type: none"> • Shape that will not collapse easily • High filtration efficiency • Good breathability • without valve • Quality complaint with standards for medical N95 respirator: <ul style="list-style-type: none"> b. NIOSHN95,EN 149, FFP2/as per BIS 9473:2002 or equivalent • Fluid resistance: minimum 80mmHg pressure based on ASTM F1862, ISO 22609, or equivalent • Quality complaint with standards for particulate respirator that can be worn with full-face shield. <p>NOTE:- Supplier will be bound to supply 75% of the ordered quantity of N-95 mask with adjustable head loop and 25% with ear loop.</p>

Note:-

- It may be noted that if any further amendments are issued then a corrigendum will be published and informed.
- Rest of the terms and conditions will remain the same.

(S. Meena)
Executive Director (Proc.)
RMSCL

W →