



## Rajasthan Medical Services Corporation Limited

Gandhi Block, SwasthyaBhawan, Tilak Marg, C-Scheme, Jaipur-302005 (Raj.)

Phone No: 0141-2228066,2228064 CIN:U24232RJ2011SGC035067

Website: http://rmsc.health.rajasthan.gov.in

E mail: edprmsc@gmail.com

Ref. No.:- F.02(96)/RMSCL/Proc/N-95 mask /NIB-07/2021/375 Dated: 02/07-12021

## Corrigendum-III

Subject: - Date extension and Amendment in specification.

Ref.:- F.02(96)/RMSCL/Proc/N-95 mask /NIB-07/2021/291 Dated: 31.05.2021

(Technical bid opening due on dated -02.07.2021)

The following amendments are made:

Last date and time of 02.07.2021 up to submission of online bids 3:00 PM	Amended Date  Last date and time of	
0. 02.07.2021 up to		
	submission of online bids	09.07.2021 up to 3:00 PM
EMD, Tender fees, RISL fees through challan and Physically 92.07.2021 up to 3:00 PM	EMD, Tender fees, RISL fees through challan and Physically	09.07.2021 up to 3:00 PM
Date and time of opening of O2.07.2021 up to Online technical bids 4:00 PM	Date and time of opening of Online technical bids	09.07.2021 up to 4:00 PM
<ul> <li>ANNEXURE - VI and in BoQ</li> <li>N-95 Mask</li> <li>Shape that will not collapse easily</li> <li>High filtration efficiency</li> <li>Good breathability</li> <li>without valve</li> <li>Quality complaint with standards for medical N95 respirator: <ul> <li>a. NIOSHN95,EN 149, FFP2/as per BIS 9473:2002 or equivalent</li> </ul> </li> <li>Fluid resistance: minimum 80mmHg pressure based on ASTM F1862, ISO 22609, or equivalent</li> <li>Quality complaint with standards for particulate respirator that can be worn with full-face shield.</li> </ul>	ANNEXURE – VI and in BoQ  N-95 Mask  Shape that will not collapse easily High filtration efficiency Good breathability without valve Quality complaint with standards for medical N95 respirator:  a. NIOSHN95,EN 149, FFP2/as per BIS 9473:2002 or equivalent Fluid resistance: minimum 80mmHg pressure based on ASTM F1862, ISO 22609, or equivalent Quality complaint with standards for particulate respirator that can be worn with full-face shield.  NOTE:- Supplier will be bound to supply 75% of the ordered quantity of N-95 mask with	

## Note:-

- It may be noted that if any further amendments are issued then a corrigendum will be published and informed.
- Rest of the terms and conditions will remain the same.

Gupta) Executive Director (Proc.) Senior Manager (P)

R.M.S.C.L., Jaipur

RMSCL