



Rajasthan Medical Services Corporation Limited

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur-302005 (Raj.)

Phone No: 0141-2228066 , 2228064 Fax No. 0141-2228065 Website: <http://rmsc.health.rajasthan.gov.in>

CIN:U24232RJ2011SGC035067

E_mail : edprmsc@gmail.com

Ref. No.:- F.02(57)/RMSCL/Proc./S&S(MD)/NIB-5/2018/ 295

Dated: 23/07/2018

Corrigendum - II

Subject:- Amendments in technical specifications, conditions and BoQ

Ref.:- NIB No. F.02(57)/RMSCL/Proc./S&S(MD)/NIB-5/2018/269 Dated:20.06.2018 and Corrigendum 287 dated 12.07.2018

S. No	Existing condition / technical specification/Packing Unit/Quantity	Amended condition / technical specification/ Packing Unit/Quantity/Date Extension
1.	<p>Clause 7(b) OPENING OF TECHNICAL BID AND TECHNICAL EVALUATION</p> <p>b) Technical Evaluation:- Technical Evaluation of the Bid will be done in two stages. (i) Technical Evaluation on the basis of documents. This is to examine whether the bidder qualifies as per given eligibility and other prescribed conditions. (ii) Evaluation / Examination / Testing of samples of Items. Bids of the item, samples not found technically fit in such evaluation will be declared as rejected / non-responsive.</p>	<p>Clause 7(b)</p> <p>b) Technical Evaluation:- Technical Evaluation of the Bid will be done on the basis of documents. This is to examine whether the bidder qualifies as per given eligibility and other prescribed conditions.</p>
2	<p>Clause 8.1 PRODUCTION OF SAMPLES</p> <p>(a) Bidder shall submit minimum 45 and maximum 50 sample units of each quoted item, free of cost. The items submitted as samples should be of the same specifications as asked for in the bid. Any deviation from specifications will result in the rejection of the sample. The samples (normal/ regular sales packs) will be used for quality evaluation by the technical evaluation committee and/ or laboratory analysis, as the case may be, as decided by the Tender inviting authority. Samples of the items which are supposed to be sterile should be sterile. The decision based on quality evaluation of the sample will be final for the purpose of this tender.</p> <p>(b) The samples for evaluation shall be submitted in a separate sealed cover superscripted by "Tender No. _____". The sample as above shall be submitted on or before the time of depositing Bid Security, or within 3 days of technical bid opening. The bidder should submit, along with the samples, the list of sample of sample items in the given format in Annexure -XI.</p>	Deleted
3	<p>Annexure XI</p> <p>PROFORMA FOR SUBMISSION OF SURGICAL / SUTURE SAMPLES</p>	Deleted

(Handwritten signature)



4

Annexure-XIV and BOQ:-

Annexure-XIV and BOQ :- The following surgical drug items in the bid are added:-

Code No.	Name of item with specification	Packing Unit	Class as per medical devices rules 2017	Tender Qty.
S-15(d)	<p>Sterile Disposable (Single Use) Teflon / PTFE I.V. Cannula with integrated 3 Way stop cock. Size 22G</p> <ul style="list-style-type: none"> • Should be packed in transparent, single blister pack. • Should conform to IS 10555 standard 	Each Piece	Class B	11546352
S-23(a)	<p>Ryle's Tube / Nasogastric Tube (P.V.C.) with radio opaque lining. Size: 10</p> <ul style="list-style-type: none"> • Soft, kink resistant PVC tubing for atraumatic intubation • Closed distal end should be coned with radio opaque material for accurate intubation • Four lateral eyes for greater efficiency • Radio opaque line • Marking at 50, 60, 70 cm from tip • Colour coded funnel • With luer connector / closure • Length 105 cm 	Each Piece	Class B	76096
S-23(b)	<p>Ryle's Tube / Nasogastric Tube (P.V.C.) with radio opaque lining. Size: 12</p> <ul style="list-style-type: none"> • Soft, kink resistant PVC tubing for atraumatic intubation • Closed distal end should be coned with 	Each Piece	Class B	114640



- radio opaque material for accurate intubation
- Four lateral eyes for greater efficiency
- Radio opaque line
- Marking at 50, 60, 70 cm from tip
- Colour coded funnel
- With luer connector / closure
- Length 105 cm

BOQ :- Three surgical drug items are being added in the bid. The BOQ is replaced with new BOQ. Please ensure that financial bid is submitted in new BOQ only.

(Total Surgical items in Bid :36)

Annexure-XIV and BOQ are to be read as follows:-
Technical Specification of S-18, S-19, S-20, S-80

5 Annexure-XIV and BOQ
Technical Specification of S-18, S-19, S-20, S-80

S-18	Paper Adhesive Plaster 1" X 9.0 mts (with cutter) Non woven adhesive tape, hypoallergic, should have some stretch bonding	Unit	Class B
S-19	Paper Adhesive Plaster 2" X 9.0 mts (with cutter) Non woven adhesive tape hypoallergic, should have some stretch bonding	Unit	Class B
S-20	Paper Adhesive Plaster 3" X 9.0 mts (with cutter) Non woven adhesive tape hypoallergic, should have some stretch bonding	Unit	Class B
S-80	Bone Wax Sterilised	2gm/ Packet	Class C

S-18	Paper Adhesive Plaster 1" X 9.0 mts (with cutter) Non woven adhesive tape, hypoallergic, should have stretch bonding	Unit	Class B
S-19	Paper Adhesive Plaster 2" X 9.0 mts (with cutter) Non woven adhesive tape hypoallergic, should have stretch bonding	Unit	Class B
S-20	Paper Adhesive Plaster 3" X 9.0 mts (with cutter) Non woven adhesive tape hypoallergic, should have stretch bonding	Unit	Class B
S-80	Bone Wax Sterilised	2.5 gm/ Packet	Class C

Note:-

- It may be noted that if any type of amendments required then further corrigendum will be published and informed.
- Rest terms and conditions will remain the same.
- It bears approval of purchase committee of RMSCL.

Handwritten initials/signature

Handwritten signature
 (V.C. Bunkar) RAcS
Executive Director (Proc.)
RMSCL

