

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, SwasthyaBhawan, TilakMarg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax: 0141-2228065

E-mail: edepmrmsc-ri@nic.in

Website: http://rmsc.health.rajasthan.gov.in

CIN:U24232RJ2011SGC035067

No. F-8(302) RMSC/EPM/M-2/NIB-542/2022-23/

Dated: 06/05/2022

All Principals Medical Colleges, All the Superintendents of Medical Colleges & Attached Hospitals. Director (PH/HA/IEC/AIDS/RCH)

Sub:- Approval/Purchase of Item, 'Single Puncture Laparoscope for Laparoscopic Sterilization with CO2 Insufflators and Cold Light Source, for the Rate Contract and supply, ending on 30.04.2024

Please find enclosed herewith approved firm and Single Puncture Laparoscope for Laparoscopic Sterilization with CO₂ Insufflators and Cold Light Source for the rate contract and supply ending on 30.04.2024 The rate contract and its terms & conditions are also available on Rajasthan Medical Services Corporation Ltd., Website www. rmsc.health.rajasthan.gov.in

In reference to the above rate contract No. F-8(302) RMSC/EPM/M-2/NIB-542/2022-23/ Dated the supplier has executed the agreement with RMSC, as enclosed herewith for your information & necessary action. This bears an approval of M.D., RMSCL, Jaipur.

Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd, Jaipur Dated: 06/05/2022

No. F-8(302) RMSC/EPM/M-2/NIB-542/2022-23/ 2 79

Copy forwarded to the following for information & necessary action:

- 1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
- 2. P.S. to ACS/Principal Secy/Secy. Medical Health & FW Deptt. Rajasthan, Jaipur.
- 3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
- 4. P.S. to Mission Director NHM Jaipur.
- 5. PA to Managing Director, RMSCL
- 6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
- 7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar& Udaipur.
- 8. Principal R.U.H.S. College of Dental Science Jaipur.
- 9. Principal, RUHS College of Medical Scinces, Jaipur.
- 10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Scinces, Jaipur).
- 11. Superintendent, SMS Hospital, Zenana Hospital, MahilaChikitsalya, Mental Hospital, J.K. Lon Hospital, T.B &
- 12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
- 13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
- 14. Addl. Director, Medical & Health Services (PW)/(RH)/(Hospital Admn.)
- 15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
- 16. All P.M.O.'s.
- 17. All CM&HO's.
- 18. All Dy. CM&HO's.
- 20. All Zonal BME are directed to carry out inspection of the supplied equipment as per RMSCL Letter no. 2513
- 21. M/s Heager India Private Limited, Plot No. 444,2nd Floor, Functional Industrial Estate, Patparganj Industrial Area, New Delhi-110092, Email: info@heagerindia.com
- 22. Guard file.

Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd, Jaipur

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, SwasthyaBhawan, TilakMarg, C-Scheme, Jaipur - 302005
Ph. No. 0141-2223887, Fax: 0141-2228065
CIN:U24232RJ2011SGC035067

E-mail: edepmrmsc-rj@nic.in
Website: http://rmsc.health.rajasthan.gov.

No. F-8(302) RMSC/EPM/M-2/NIB-542/2022-23/ 2-80

E-mail: edepmrmsc-rj@nic.in / Website: http://rmsc.health.rajasthan.gov.in Dated: 06/05/2022

Rate Contract and Price Details

S.N.		Item Detail				
1.	Name of Equipment / Item		Single Puncture Laparoscope for Laparoscopic Sterilization			
			with CO ₂ Insufflators and Cold Light Source			
2.	Name & Address of Firm		M/s Heager India Private Limited			
			Plot No. 444,2nd Floor,			
		Charles No.	Functional Industrial Estate,	*		
			Patparganj Industrial Area, New Delhi-110092,			
			Email: info@heagerindia.com	-		
3.	Bidder S	The state of the s				
4.	Performa	ance Security	Performance Security Deposit by Bank			
			027GT02221120001 Dated 22.04.2022 (HDFC Ba			
	F	120, Patparganj Industrial Area, New Delhi-110092 for Amount of				
			Rs. 10,15,000/- (Rupees Ten Lakh Fifteen Thousand	d Only) (Valid Up		
			to 31.10.2027)			
5.	HSN Co	de	90185090			
6.	Brand/ N	Model .	HEAGER			
	Make		•			
7.	Packing	Unit	Each			
8.		& Validity	R.C. No. 761 (Valid Up to 30.04.2024)			
9.	Indicativ	e Quantity 50				
10.	Rates for supply to RMSC from Suppliers					
	(i) Bas		7,25,000.00			
		(ii) IGST @ 12% (in Rs.)				
		(iii) CGST @ 6% (in Rs.)				
	(iv) SGS	43500.00				
11	(v) App	812000.00				
11.	Rate For Supply to Consignee From RMSC (i) Basic Rate (in Rs.) 7,25,			7,25,000.00		
	(ii) RM	36250.00				
	(iii) CG	45675.00				
	(iv) SG	45675.00				
		852600.00				
12.	(v) Approved Rate for Consignee from RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.) 852600.00 SPECIFICATIONS OF Single Puncture Laparoscope for Laparoscopic Sterilization					
	with CO ₂ Insufflators and Cold Light Source					
	S.N.					
	1.	Telegopes Telegope 0 degree with nevellel/straight are piece 10 to 12 mm dismeter Should				
	1.	Telescope: Telescope 0 degree with parallel/straight eye piece, 10 to 12 mm diameter. Should have 6 mm instrument channel built-in ring applicator Working length of 270-275 mm/10mm 0 degree telescope. Fibre optic light trans. mission incorporated, should be compatible with the commonly available light cables (necessary adaptors should be provided) Can be sterilized by				
*		degree telescope. I commonly available	Fibre optic light trans. mission incorporated, should be c	ompa		



भुस्यमंत्री नि:शुल्क जांच योजन्

T	2.	Trocar & Cannula: Cannula size +1 mm more than the telescope diameter; should have an
		automatic valve and stopcock for insufflations. (Compatible with supplied telescope). Trocar
		should have pyramidal tip. Tip should be sharp enough but should not injure the viscera. The
		religiii of the frocar should be 1/5 mm±10 mm and the length of the cannula should be 100.110.
		min. (mina abdominal part)
	3.	Ring Applicator: Ring applicator for use with parallel/straight eyepiece telescope compatible
	,	will the above telescope. Capable of loading two silastic rings. The ring applicator has to be
		rany dismantable into different parts like prone, inner tube, outer tube, knurled ring etc.
	4.	Cone: Suitable cones for loading rings to the above applicator made of Teflon / silicon / SS
		317 grade
	5.	Slide/guide: Suitable slide/guide for loading rings to the above applicator.
	6.	Veress Needle: Veress Needle with spring loaded blunt stylet with lucr lock and length of 10 cm
	7.	Veress Needle: Veress Needle with spring loaded blunt stylet with lucr lock and length of 15
		Cit
	8.	Essential Spares:
		(i) Spare washers: Spare washers for trocar, Cannula and automatic valve
		(II) Kits for cleaning: Trocar Brush, Cannula Brush, Silicon Oil 50ml Special Judgicent for
11		stopcock & telescope lens cleaner.
		(iii) Spare parts: Spring of ring applicator, finger ring (thumb) for ring applicator, knurled
		solew for ring applicator, inner sheath of ring applicator, tension rod with grasner (proper insert)
		for this applicator, adaptor for fiber optic light cable for telescope, stoncock for cappula cos
П		miet, spring cap for stopcock
11	0	(iii) Manuals: Operational & service manuals
П	9.	Supplier should provide a plastic storage brief case with foaming inside for laparoscope
11	10	relescope and all hand instruments and accessories for storage & transportation
	10.	Environment: Telescope should be specified for operation up to 45° C and 100% relative
11	11	numidity
	11.	Electronic CO2 insufflators with pin index connection. Should have an adjustable
		minimum now rate of 0 to 40 ftr Per minute and a pressure range adjustable between 0.20
П		min fig. Fressure and flow rate should be display on the front panel Drovided with all and
П		autociavable tubing with luer attachment, Instrument should work on a supply of 190, 240
		v. with a frequency of 50 Hz single phase. Fit for purpose. The unit should be comply with
		relevant LEC salety standards.
Ш		Automatic gas release system for constant monitoring of intra abdominal pressure and checking
П		over pressure with automatic back release of CO ₂ gas within 5 seconds should include 1 pack/10
Π		filter for CO ₂ gas.
		1. The machine should give an audible alarm signal in case of wrongly placed veress
П		pneumoperitonium needle and sudden block in the CO ₂ flow from machine
		2. The insufflator should also give audible alarm in case of overpressure and release of it automatically
П		
		3. Both the preset value and actual value for pressure and flow should be displayed at the same time on the front panel of the machine
1	12.	High Pressure hose High Pressure have the life in the
	14.	High Pressure hose: High Pressure hose suitable to connect the insufflators with pin indexed CO2 cylinder.
止	13.	
		Wrench kit: Suitable for connecting the insufflators to CO ₂ cylinder
	14.	Carbon Diovide Rottle (avlindar): 5 V. a constitution of the Carbon Diovide Rottle (avlindar):
		Carbon Dioxide Bottle (cylinder): 5 Kg. or equivalent Capacity Carbon Dioxide bottle with pin
		index connection. The unit should automatically display status of the CO ₂ conditioning by it
		electronic Bar Graph / Digital display on the front panel of the insufflator. The unit should have
		inbuilt pressure regulator to detect high & low pressure of CO ₂ gas automatically at the input.
	15.	Cold Light Sources: I ED Light course with any 14 11 14
		Cold Light Sources: LED Light source with one / two light outlets, having power more than
		100 Watts, and average LED lamp life of minimum 25000 hrs. or More. Electrical Safety
		Certification IEC- 601-1(test report must be submitted in technical bid) Rechargeable Handy LED Back Up light source compatible with SP leaseness in the light source compatible with the light source compatible with SP leaseness in the light source compatible with the light source compatible w
		Handy LED Back Up light source compatible with SP laparoscope is also provided. Handy LED battery light source (Rechargeable with Charges) which can be directly
		battery light source (Rechargeable with Charger) which can be directly mounted on to the laparoscope Telescope without any delay in case of break down with Main light source type. The
	,	The

my

		unit should comply with relevant IEC safety standards.
	16.	Fibre Optic Light cable: 4.5-10.0 mm in diameter. 250-300cm in length compatible with the
		cold light source and can be fitted the commonly available telescope (necessary adaptors may be
		provided) provided
	17.	UPS: UPS-1.0 KVA with one hour backup time, good quality, ISI/CE approved.
	18.	Essential Parts:
	,	(i) Mains cord: Compatible with insufflators and cold light source.
		(ii) Manuals: Operation & service manual (for insufflators & Light source)
	19.	Tormann Champer (for Sterilization of Languages)
		Dimensions of Formalin chambers 65 cm (±10cm) X 20 cm (±2cm) X 20 cm (±2cm) with three tray made of white Opaque Acmilia
		1 OF WILLO ODAULE ACIVILE
	20.	Tray for sterilization of Langroscopes. Dimensions of outer statistics.
		1 - Star A 13 VIII + Z VIII I VIIIIMENTO CONSCITO III IIIIA I O III.
		and the state of t
	21.	Compatibility & Core items: The core laparoscopy items such as Telescope, Trocars &
		Table 1 Dulle 1 Dulle 1 Dulle 1 De 1
		The state of the s
		TOTAL TIPO TOTAL
	00	The state of the s
	22.	Core items/products should be US FDA/European CE(notified hody)/DIC and a second should be used to
	22	monetoned safety standard in specifications.
	23.	Guarantee: Three years on equipment from the date of installation.
	24.	CMC: CMC will be given @ 3 % of net rate excluding GST (as applicable). The CMC
	25.	white does not have years to really passes after completion of Champion of Cha
	25.	itates of Consumable/spare parts: Rates of consumable and spare and an area
	26.	The oderbrugge product of broaden by bidger in B()()
	27.	Installation will be done by supplier free of cost.
	28.	The service engineer should be based in Rajasthan.
	29.	Demonstration of Equipment is must for final approval.
	29.	The company should mention the make & model name/number of the quoted equipment and
		and the following of the duoted model in the Technical hid along with
	30.	
	30.	Firm should mention all the pre-installation requirements in technical bid.
		(1) Bidders are requested to send with bid, printed descriptive literature of the quoted items
		(11) Il Bidder supplied to or have rate contract of quoted items with any other of
		wall one year, he may be asked to provide copies of purchase orders invoices and make I
		401/41.600.
•	The sore	gement signed between DMSCI and discourse

• The agreement signed between RMSCL and the firm forms integral part of the rate contract.

Executive Director (EPM)
RMSC, Jaipur

Dated: 06/05/2022

No. F-8(302) RMSC/EPM/M-2/NIB-542/2022-23/2-80

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.

2. P.S. to ACS/Principal Secy/Secy. Medical Health & FW Deptt. Rajasthan, Jaipur.

3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur

4. P.S. to Mission Director NHM Jaipur.5. PA to Managing Director, RMSCL

6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).

7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar& Udaipur.

Principal R.U.H.S. College of Dental Science Jaipur.
 Principal, RUHS College of Medical Scinces, Jaipur.

10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Scinces, Jaipur).

मुखमंत्री नि:शुल्क जांच योज

- 11. Superintendent, SMS Hospital, Zenana Hospital, MahilaChikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
- 12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital-Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
- 13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
- 14. Addl. Director, Medical & Health Services (PW)/ (RH)/(Hospital Admn.)
- 15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota &Bharatpur.
- 16. All P.M.O.'s.
- 17. All CM&HO's.
- 18. All Dy. CM&HO's.
- 19. AGM (IT), RMSC, Jaipur.
- 20. All Zonal BME are directed to carry out inspection of the supplied equipment as per RMSCL Letter no. 2513 dated 10-07-2018.
- 21. M/s Heager India Private Limited, Plot No. 444,2nd Floor, Functional Industrial Estate, Patparganj Industrial Area, New Delhi-110092, Email: info@heagerindia.com

22. Guard file.

Executive Director (EPM)
RMSC, Jaipur