

Rajasthan Medical Services Corporation Limited, Jaipur

'D'-block, swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur

Phone no. 0141-2223887, Fax no. 0141-2228065

E-mail:edepmrmsc-rj@nic.in

CIN: U24232RJ2011SGC035067

Website: http://rmsc.health.rajasthan.gov.in

No. F-8() RMSC/EPM/M-5/NIB-795/2023-24/96 &

Dated | o | 1 | re

All Principals Medical Colleges,

All the Superintendents of Medical Colleges & Attached Hospitals.

Director (PH/HA/IEC/AIDS/RCH)

Sub:- Approval/Purchase of Item, 'Radiant Warmer, for the Rate Contract and supply, ending on 31.12.2025.

Please find enclosed herewith approved firm and Radiant Warmer for the Rate Contract and supply, ending on 31.12.2025. The rate contract and its terms & conditions are also available on Rajasthan Medical Services Corporation Ltd., Website www.rmsc.health.rajasthan.gov.in

In reference to the above rate contract no. F-8() RMSC/EPM/M-5/NIB-795/2023-24/.....dated...... the supplier has executed the agreement with RMSC, as enclosed herewith for your information & necessary action. This bears an approval of M.D., RMSCL, and Jaipur.

Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd, Jaipur

No. F-8()RMSC/EPM/M-5/NIB-795/2023-24/968

Copy forwarded to the following for information & necessary action:

- 1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
- 2. P.S. to ACS/ PHS/ Secy., Medical Health & FW Deptt. Rajasthan, Jaipur.
- 3. PS to ACS/PHS/Secy., Medical Education Department, Rajasthan, Jaipur
- 4. P.S. to Mission Director NHM Jaipur.
- 5. PA to Managing Director, RMSCL
- 6. Director, (PH) /(Aids)/(FW)/(RCH)/(IEC)/(ESI).
- 7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
- 8. Principal R.U.H.S. College of Dental Science Jaipur.
- 9. Principal, RUHS College of Medical Sciences, Jaipur.
- 10. Superintendent Jaipuria Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
- 11. Superintendent, SMS Hospital, Zenana Hospital, Mahila Chikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
- 12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital-Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
- 13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
- 14. Addl. Director, Medical & Health Services (PW)/ (RH)/(Hospital Admin.)
- 15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
- 16. All CM&HO's., All P.M.O.'s. and All Dy. CM&HO's, Rajasthan.
- 17. Joint Director (IT), for uploading in e-upkaran software and RMSCL website both.
- 18. All BME, RMSC
- 19. M/s Nice Neotech Medical system Pvt. Ltd., No. 85, Krishna Industrial Estate, Vanagaram, Mattukuppam, Chennai-600095Email-info@nicetech.com, marketing@niceneotech.com tender@niceneotech.com

20. Guard file.

Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd, Jaipur



Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, SwasthyaBhawan, TilakMarg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax: 0141-2228065

CIN:U24232RJ2011SGC035067

E-mail: edepmrmsc-rj@nic.in

Website: http://rmsc.health.rajasthan.gov.in

No. F-80RMSC/EPM/M-5//NIB-795/2023-24/968

Dated 10/1/24

Rate Contract and Price Details

S.N.	Item	Detail
1.	Name of Equipment / Item	Radiant Warmer
2.	Name & Address of Firm	M/s Nice Neotech Medical system Pvt. Ltd., No. 85, Krishna Industrial Estate, Vanagaram, Mattukuppam, Chennai-600095 Email-info@nicetech.com, marketing@niceneotech.com tender@niceneotech.com
3.	Bidder Status	Manufacturer
4.	Performance Security	BG NO. 0149123IPG000027 Date 13.12.2023 for Rs. 49,09,800.00 Indian Bank Limited., Plot No. 254-260, Avvai shanmugham salai, Royapettah, Chennai-600014 Email: valasaravakkam@indianbank.co.in
5.	Make/Model	NICE Neotech/ Nice2010 BC
	HSN Code	9018
6.	Packing Unit	Each
7.	R/C No. & Validity	RC No. 833 Valid Up to 31.12.2025
8.	Indicative Quantity	1500
9.	Rates for supply to RMSC from Suppliers	
	(i) Basic Rate (in Rs.)	58450.00
	(ii) IGST @ 12% (in Rs.)	7014.00
	(iii) CGST (in Rs.)	0.00
	(iv) SGST(in Rs.)	0.00
	(v) Approved Rate From {(i)+(ii)+(iii)+(iv)} (in Rs.)	Supplier To RMSC 65464.00
10.	Rate For Supply to Consignee From RMSC	
	(i) Basic Rate (in Rs.)	58450.00
	(ii) RMSC Surcharge@5% Or	n (i) (in Rs.) 2922.5
	(iii) CGST@ 6% { On (i)+(ii)}	(in Rs.) 3682.35
	(iv) SGST@ 6% { On (i)+(ii)}	(in Rs.) 3682.35
	(v) Approved Rate for Consign	nee from RMSC 68737.20
	$\{(i)+(ii)+(iii)+(iv)\}\ (in\ Rs.)$	
11.	Technical Specifications of Radiant Warmer	
	 Essential parts a) Quartz/Ceramic/Calrod based warming system with microprocessor based controls, probes & alarms and has servo and manual temperature control facility. b) Cart & bassinet c) Examination light 	





- d) X-ray cassette holder
- e) Drawer
- f) IV stand (SS)
- g) Monitor shelve(Minimum 10X12 Inch)
- Should have two sensing probes to monitor Core and peripheral temperatures
- 3. It should have thermistor probes to improve accuracy of temperature readings.
- 4. Once the skin probe gets off from the bed of the patient for more than 10 minutes

To avoid overheating of patient, heater output should automatically cut off to 20-25% if the heater output is more than 60% and should alarm continuously for check sensor.

- 5. Cart
- a) Should swivel on 4 wheels of at least 4" diameter with foot operated brakes on at least 2 front wheels
- 6. Dimensions-

Bed should be about 80-100 cms from the floor and 80-90 cms from the heat source, as per international guidelines.

- Bassinet
- a) Should be supplied with good quality thermal sealed & water proof Mattress.

Mattress Mattress size should be minimum 20"X 30".

Mattress should have density approx. 20-25 kg/m3.

- b) Transparent collapsible unbreakable polycarbonate side walls easily detachable for cleaning Mattress.
- 8 Bassinet tilt
- a) Should allow tilt for Trendelenburg as well as reverse Trendelenburg position
- b) Should have continuous jerk free variable bed tilting mechanism for a bed tilt on either side to ±10-15 degree
- Warming system
- a) Quartz/Ceramic/Calrod based heating system
- b) Control Microprocessor controlled with soft touch control Panel
- c) Self test function performed at power on
- d) Digital display should show following parameters
- i. Set temperature
- ii. Present temperature of the baby
- iii. Heater output
- e) Mode Manual, pre heat/pre warm & skin (servo)
- f) Manual mode
- i. Adjustable in steps from 0 to 100% in increments of 10%
- ii. Heater power should be reduced to 50 60% after 10-15minutes in manual mode for baby safety
- g) Skin mode (servo mode)
- i. Set point range 32 to 38 degrees C
- h) Skin temperature display
- i. Accuracy ±0.1-0.3 degrees C
- ii. Resolution 0.1 degree C
- i) Temperature probe-Only thermistor probes and wire should be easy to clean and long lasting
- j) No need of temperature probe calibration
- k) Control unit should have facility to convert Centigrade to Fahrenheit conversion
- 1) Continuous measurement with Large easy to read display
- 10. Alarms
- a) Audiovisual alarms-
- i. Probe failure
- ii. Heater failure
- iii. High and low baby temperature (more than 0.5 deg Cdifference)
- iv. Power failure
- v. System failure
- vi. Silence/reset switch

Fami



- 11. Heating unit Should be swivable (if required) for accommodating X-Ray unit and should have self lock facility.
- 12. LED observation light should be provided for observing of baby.
- 13. Apgar Timer
- a) Timer with stopwatch facility
- b) Reset facility
- 14. I.V. stand
- a) Strong IV stand (S.S) with height adjustable and facility to fixlarge number of infusion pumps
- 15. Integrated monitor shelve.
- 16. Should have integrated atleast one large size drawer to keep articles
- 17. Integrated X- Ray cassette holder
- a) Sliding holder located just below under surface of bassinet, with markings to help placement of cassette. X ray cassette tray should accept up to 20 mm thick xray cassette.
- 18. Power consumption Less than 1 K.W
- 19. All metal parts of the equipment should be corrosion resistant and Epoxy/Powder coated
- 20. All consumables required for installation and standardization of system to be given free of cost
- 21. Availability of spares for at least 7 years after date of installation
- 22. Standards, safety and training
- a) Should be US FDA or European CE/BIS approved product
- b) Manufacturer should be ISO 13485 certified for quality standards.
- c) Electrical safety conforms to standards for electrical safety IEC 60601-1 (OR EQUIVALENT international/national standard)General requirement for Electrical safety of Medical Equipment, related document shall submitted in technical bid.
- d) Should have local service facility. The service provider should have the necessary equipments recommended by the manufacturer to carry out preventive maintenance test as per guidelines provided in the service/maintenance manual
- 23. Warranty 3 years from date of installation.
- 24. Power supply
- a) Power input to be 220-240VAC, 50Hz
- b) Suitable Auto voltage corrector with spike protector should be available
- 25. Environmental factors
- a) Shall meet IEC-60601-1-2 :2001 General Requirements of Safety for Electromagnetic Compatibility or should comply with 89/366/EEC; EMC-directive
- b) The unit shall be capable of being stored continuously in ambient temperature of 0-50deg C and relative humidity of 15- 90%
- c) The unit shall be capable of operating continuously in ambient temperature of 20-40 deg C and relative humidity of 15-90%
- 26. Documentation
- a) User/Technical/Maintenance manuals to be supplied in English/Hindi
- b) Certificate of calibration and inspection from factory
- c) List of important spares and accessories with their part number and costing
- 27. Essential accessories to be supplied at initial purchase with each piece of equipment
- a) Reusable temperature probes (For measuring core temperature): 02 (Two) per radiant warmer Or 200 disposable probes per radiant warmer
- b) Reusable thermistor temperature probes (For measuringperiphery temperature): 02(Two) per radiant warmer.
- 28. The rates of consumable accessories should also be quoted separately
 Onsite physical demonstration/training of the equipment to all the endusers with all the requested facilities will be mandatory
- The agreement signed between RMSCL and the firm forms integral part of the rate contract.

Banin

Executive Director (EPM)
RMSC, Jaipur



No. F-80RMSC/EPM/M-5/NIB-795/2023-24/968

Dated: 10(01/2024

Copy forwarded to the following for information & necessary action:

- 1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
- 2. P.S. to ACS/PHS/Secy., Medical Health & FW Deptt. Rajasthan, Jaipur.
- 3. PS to ACS/PHS/Secy., Medical Education Department, Rajasthan, Jaipur
- 4. P.S. to Mission Director NHM Jaipur.
- 5. PA to Managing Director, RMSCL
- 6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
- 7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar& Udaipur.
- 8. Principal R.U.H.S. College of Dental Science Jaipur.
- 9. Principal, RUHS College of Medical Scinces, Jaipur.
- 10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Scinces, Jaipur).
- 11. Superintendent, SMS Hospital, Zenana Hospital, Mahila Chikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
- 12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
- 13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
- 14. Addl. Director, Medical & Health Services (PW)/ (RH)/(Hospital Admn.)
- 15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
- 16. All CM&HO's., All P.M.O.'s. and All Dy. CM&HO's, Rajasthan.
- 17. Joint Director (IT), for uploading in e-tmts software and RMSCL website both.
- 18. All BME, RMSC
- 19. M/s Nice Neotech Medical system Pvt. Ltd., No. 85, Krishna Industrial Estate, Vanagaram, Mattukuppara, Chennai-600095, Email-info@nicetech.com, marketing@niceneotech.com, tender@niceneotech.com

20. Guard file.

Executive Director (EPM)

RMSC, Jaipur