



RAJASTHAN MEDICAL SERVICES CORPORATION LTD.

D- Block, SwasthyaBhawan, TilakMarg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN: U24232RJ2011SGC035067

E-Mail -edepmrmsc-ri@nic.in

No. F-8(RMSC/EPM/M-7/2020-21/NIB-613/ 1133

Dated : 29/09/2021

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: Amended Technical Specification, and Time extension of item UV water purifier (50 LPH) of NIB No.F-8()RMSC/EPM/SM-3/2021-22/NIB-613/965 dated 12.08.2021

In reference to subject cited above NIB-613, various representation received from the firms and issues raised by the bidder are examined by the competent authorities and technical committee. The following clarification/corrigendum/addendum is issued for inclusion in bid document of items as below:-

A-Revised Technical Specifications of UV water purifier (50 LPH)**Newly added:-**

- 500 Ltr. Storage tank of food grade plastic material for storage drinking water with water tap along with suitable stand for the tank has to be provided .
- Height of stand for 500 ltr. storage tank should not be less than 3.5 feet and stand be made up of iron with rust proof coating.
- Rates of 500 ltr. storage tank of food grade plastic material and the suitable stand for 500 ltr. storage tank shall be provided in separate sealed envelope along with technical bid.

L1-Bidder:- L1 bidder will be decided on the basis of L1 rate received for UV water purifier only as mention in BF-4 and online BOQ.

B-Revised bid schedule:-

- E-bids are invited as per following revised time schedule:-

Existing Dates			Extended Dates		
Last date & time for sale of bid form	Last date & time of receipt of bid form	Date & time of opening of technical bid	Last date & time for sale of bid form	Last date & time of receipt of bid form	Date & time of opening of technical bid
30.09.2021 06:00 p.m.	30.09.2021 06:00 p.m.	1.10.2021 03:00 p.m.	11.10.2021 06:00 p.m.	11.10.2021 06:00 p.m.	12.10.2021 03:00 p.m.

It is also clarified that information of award of contract shall be communicated to all participating bidders on the website www.rmsc.nic.in and sppp.raj.nic.in. Please note that individual bidder will not be intimated."

Note:-Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-V, Schedule of Supply, and Point no. 3) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

This bears the approval of M.D., RMSCL, Jaipur.

Executive Director (EPM)
RMSCL, Jaipur