



RAJASTHAN MEDICAL SERVICES CORPORATION LTD.
D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065

E-Mail - edepmrmsc-rj@nic.in

No. F-8() RMSC/EPM/M-4/NIB-801/2023-24/ 127

Date: 10-05-2024

Office Order

List of bidders declared responsive/non responsive as per documentary & technical evaluation in bid number NIB No. F-8() RMSC/EPM/M-4/NIB-801/2023-24/427 Date: 25.08.2023 for item (1) LED O.T. LIGHT DOUBLE DOME and (2) LED O.T. LIGHT CELLING SINGLE DOME technical bid opening date being 19.10.2023 and as per decision taken in the purchase committee dated 06.05.2024 headed by M.D., RMSCL is under.

S.No	Firm Name	Quoted Item	Reason for Responsive/ Non Responsive	Status
1	2	3	4	5
1.	M/s Denis Medicare Pvt Ltd	LED OT LIGHT CELLING SINGLE DOME	Technically Approved in Demo	Responsive
		LED OT LIGHT DOUBLE DOME	Past performance not complied	Non-Responsive
3.	M/s SI SURGICAL CORPORATION	LED OT LIGHT CELLING SINGLE DOME	As per tender specification - Color Temp Variable (3800 K – 4500 K) (Report from NABL Lab/ Govt. approved laboratories WIZ-STQC/ITDL or from any international certifying agency under notified body. Certificate should be attached in technical bid.)for LED OT light single Dome. Firm submitted test report from Bay Area Compliance Labortaories Corp. (Dongguan), China. Not complied.	Non-Responsive
		And LED OT LIGHT DOUBLE DOME	Color Temp Variable (3800 K – 4500 K) with CRI greater than 90 (Report from NABL Lab/ Govt. approved laboratories WIZ-STQC/ITDL or from any international certifying agency under notified body. Certificate should be attached in technical bid.) for LED OT light Double Dome. Firm submitted test report from Bay Area Compliance Labortaories Corp. (Dongguan), China. This lab is not a notified lab.	Non-Responsive
5.	M/s Surgident India	LED OT LIGHT CELLING SINGLE DOME	As per tender specification - Color Temp Variable (3800 K – 4500 K) (Report from NABL Lab/ Govt. approved laboratories WIZ-STQC/ITDL or from any international certifying agency under notified body. Certificate should be attached in technical bid.) for LED	Non-Responsive

RajKaj Ref 7105304

**Signature valid**

Digitally signed by Dr. Akash Alha
Designation: Executive Director
Date: 2024.05.10 12:43:25 IST
Reason: Approved



RAJASTHAN MEDICAL SERVICES CORPORATION LTD.
D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065

E-Mail –edepmrmc-rj@nic.in

S.No	Firm Name	Quoted Item	Reason for Responsive/ Non Responsive	Status
1	2	3	4	5
			OT light single Dome. Firm submitted these parameters on letterhead.	
		And LED OT LIGHT DOUBLE DOME	Color Temp Variable (3800 K – 4500 K) with CRI greater than 90 (Report from NABL Lab/ Govt. approved laboratories WIZ-STQC/ITDL or from any international certifying agency under notified body. Certificate should be attached in technical bid.) For LED OT light Double Dome. Firm submitted these parameters on letterhead.	Non-Responsive
7.	M/s Unisearch Medicare System Pvt. Ltd	LED OT LIGHT CELLING SINGLE DOME	Past performance not complied. Submitted new BF-7	Non-Responsive
		And LED OT LIGHT DOUBLE DOME		Non-Responsive
9.	M/s United Surgical Industries	LED OT LIGHT CELLING SINGLE DOME	As per tender specification - Color Temp Variable (3800 K – 4500 K) (Report from NABL Lab/ Govt. approved laboratories WIZ-STQC/ITDL or from any international certifying agency under notified body. Certificate should be attached in technical bid.) for LED OT light single Dome. Report not submitted	Non-Responsive
		And LED OT LIGHT DOUBLE DOME	Color Temp Variable (3800 K – 4500 K) with CRI greater than 90 (Report from NABL Lab/ Govt. approved laboratories WIZ-STQC/ITDL or from any international certifying agency under notified body. Certificate should be attached in technical bid.) for LED OT light Double Dome. Report not submitted	Non-Responsive
11.	M/s Vandana Surgicals	LED OT LIGHT CELLING SINGLE DOME	Past performance not complied because installation reports not sealed and signed by consignee.	Non-Responsive
		And LED OT LIGHT DOUBLE DOME		Non-Responsive

Online financial bids of responsive bidders for above item will be opened on 13.05.2024 At 03:00 P.M. **RajRaj** 7105384

Signature valid


Digitally signed by Dr. Akash Alha
Designation: Executive Director
Date: 2024.05.10 12:43:25 IST
Reason: Approved

LABEL FOR BIO-MEDICAL WASTE BAGS



HANDLE WITH CARE

CYTOTOXIC HAZARD SYMBOL



HANDLE WITH CARE

Annexure 'B'

LABEL FOR TRANSPORTING BIO-MEDICAL WASTE BAGS

Day Month
Year
Date of generation

Waste category Number
Waste quantity.....
Sender's Name and Address
Phone Number

Receiver's Name and Address:
Phone Number

Fax Number
Contact Person.....
In case of emergency please contact :
Name and Address :
Phone No.

Fax Number
Contact Person

Note :Label shall be non-washable and prominently visible.