

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax: 0141-2228065 CIN:U24232RJ2011SGC035067

E-mail : edepmrmsc-rj@nic.in Website: http://rmsc.health.rajasthan.gov.in

No. F-8() RMSC/EPM/M-3/21-22/NIB-635/13/6

Dated: |6/11/2021

CLARIFICATION/CORRIGENDUM/ADDENDUM

Sub:- Revised Technical specifications of the rate contract for Pulse Oxymeter (Finger Tip) for under NIB No. F-8() RMSC/EPM/M-3/21-22/NIB-635/1184 dated 12.10.2021.

In Reference to subject cited above and NIB-635, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Clarification/Corrigendum/Addendum is issued:-

Revised Technical specifications of Pulse Oximeter (Fingertip)

- 1. This fingertip oxymeter is intended for measuring the pulse rate and functional oxygen saturation (SpO₂) through a patient's finger from adult to pediatric patient.
 - 2. It should display SpO₂, PR, SpO₂ wave & plethysmogram.
 - 3. SpO₂ probe should have shielded design to block all external ambient light.
 - 4. Best Features: Auto Power Off, Water Resistant, Buzzer- Visual Alarm, Dual color OLED display/seven segment LED. Low battery indication.
 - 5. Quick, Accurate and Easy To Use It's newly designed monitor gives instant readings of blood oxygen saturation (SpO₂), pulse rate (PR), Plethysmogram and displays them on a large screen. Wavelength display clearly shows inconsistencies. Its display can rotate in four directions.
 - 6. The quoted model should be European CE (Notified Body)/USFDA/BIS Certified.

Type of Product	Digital Fingertip Pulse Oxymeter
SpO ₂ Measurement Range	70%-100%
Pulsé Measurement Range	30 - 250bpm
SpO ₂ Measurement Accuracy	70% - 100% ± 2 %
Pulse Measurement Accuracy	30 - 99bpm ± 2bpm 100-250bpm ± 2%
Pulse Indication	0.2 - 1.0% ±0.2digits; 1.1-20.0% ±20%
Battery Type	2*AA/AAA 1.5V Alkaline Batteries Type
Warranty	2 Year

1. Bidders are requested to send with bid, printed descriptive literature (catalogue) of the quoted goods.

Note:- Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-VIII) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

This bears an approval of Managing Director, Rajasthan Medical Services Corporation Limited, Jaipur.

Executive Director (EPM)
RMSCL, Jaipur