

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065
CIN:U24232RJ2011SGC035067

E-mail : edepmrmisc-rj@nic.in

Website: http://rmisc.health.rajasthan.gov.in

No. F-8() RMSC/EPM/M-6/2021-22/NIB-530/1702

Dated: 22/02/2022

OFFICE ORDER

List of bidders declared responsive/non responsive as per documentary evaluation & technical evaluation of Item **Radiology High Frequency X-Ray Machine (100mA, 300mA, 500mA and 100mA Mobile)** bid invited by No. F-8() RMSC/EPM/M-7/NIB-530/2021-22/119 Dated: 15.10.2021. The list of bidders declared responsive/ non responsive is as under -

S.N.	Firm Name	Quoted Item	Responsive/ Non-Responsive	Remarks
1	2	3	4	5
1	Allengers Medical Systems Pvt. Ltd. 205 Second Floor, Ansals Classique Tower, J Block Rajoori Garden, New Delhi email. govt@allengers.net tenders@allengers.net	100mA, 300mA, 500mA, 100mA Mobile	Responsive	Responsive for 100mA, 300mA, 500mA, 100mA Mobile
2	American Meditek Systems Pvt. Ltd. P.no.- 194, 1 st floor , HPSIDC, Industrial area, Baddi, Distt. Solan (HP) 173205 E-mail:- sales@americanmeditek.com tenders@americanmeditek.com	300mA, 500mA, 100mA Mobile	Responsive Non-Responsive for 300mA/500mA,	Responsive for 100mA Mobile, Technically (Non-Responsive) for 300mA, 500mA • only 4(Four) way table movement available • Height adjustment facility not available
3	EPSILON HEALTHCARE SOLUTIONS Pvt. Ltd. 103,104 Amer Vijay, complex SC Road Jaipur-302001 email. marketing@epsilonhealthcare.com	100mA, 300mA, 500mA, 100mA Mobile	Responsive	Responsive for 100mA, 300mA, 500mA, 100mA Mobile
4	Medion Healthcare Pvt. Ltd. P.no- K49, Additional MIDC, Anand Nagar, Ambernath (East), Thane District E-mail:- sales@medion.co.in	300mA, 500mA	(Non-Responsive) Mandatory document ISO 9001:2015 not submitted in bid.	(Non-Responsive)
5	Prognosys Medical Systems Pvt. Ltd. 168/1 Machohalli, Dasanapura Hobli, Next to vani Vidhyalaya, Off, Magadi Road, Bangalore-560091 E-mail:- info@prognosysmedical.com	500mA	(Non-Responsive) ISO13485:2016 Not Valid on bid submission date	(Non-Responsive)
6	TRIVITRON HEALTHCARE Pvt. Ltd. 15, IV Street, Abhiramapuram, Chennai-600018, India E-mail:- corporate@trivitron.com	100mA, 300mA, 500mA	Responsive	Responsive for 100mA, 300mA, 500mA

Online financial bids of responsive bidders for above item will be open on dated 25-02-2022 at 03:00 pm. Demonstration report can be obtained from office during office time.

22/2/22
Executive Director (EPM)
RMSCL, Jaipur