



**Rajasthan Medical Services Corporation Limited, Jaipur  
D-Block, Swasthya Bhawan, C-Scheme, Jaipur - 302005**

Ph. No. 0141-2223887, Fax No. 0141-2228065  
CIN : U24232RJ2011SGC035067

E-Mail – mdrmsc@nic.in; edepmrmrsc-rj@nic.in  
Website: <http://rmsc.health.rajasthan.gov.in>

F-8(189) RMSC/EPM/M-2/NIB-453/2019-20/ 3441 Dated: 30/01/2020

**Clarification/Corrigendum/Addendum**

**Subject:-Revised Technical Specifications of NIB No. F-8(189)RMSC/EPM/M-2/NIB-453/2019-20/3039 dated 9.12.2019 for item Single Channel ECG Machine, 3 Channel ECG Machine & 12 Channel ECG Machine.**

In reference to subject cited above of NIB-453, various representations received from the firms and issues raised by the bidders are being examined authorities and technical committee. The following corrigendum/addendum is issued for inclusion in bid document as below:-

**1. Revised Technical Specification:-**

**Single Channel ECG Machine-**

**Additional point:-**

1. 5 inches color or more color TFT/LCD screen.

**3 Channel ECG Machine-**

Sr. No	Existing Technical Specification	Amended Technical Specification
1.	3 channel ECG Machine with LCD display, Size 4" or more.	3 Channel ECG machine with colour TFT/LCD display, Size 5" or more.
2.	Inbuilt Rechargeable battery for at least 2 hours.	Inbuilt Rechargeable battery for at least 4 hours.
3	Frequency response: 0.05-150 Hz.	Frequency Response of 0.05-250 Hz
4	CE (European) or USFDA approved.	European CE (notified body) or USFDA approved.

**12 Channel ECG Machine-**

Sr. No	Existing Technical Specification	Amended Technical Specification
1.	12 Lead Monitoring Simultaneously on LCD display, size 6" or more	12 lead monitoring simultaneously on coloured TFT/LCD display size 8" or more
2.	Full QWERTY Alpha Numeric Key Board or Touch Screen Keyboard.	Full QWERTY Alpha Numeric Key Board with waterproof keyboard/touch screen keyboard (Touchpad).

Handwritten signature in blue ink.

Handwritten mark in blue ink, possibly '1/2'.



3	Product should be European CE or USFDA approved. (Valid certificate should be submitted in technical bid).	Product should be European CE (notified body) or USFDA approved. (Valid certificate should be submitted in technical bid).
---	--	--

**Additional point-**

1. Should have ECG acquisition of 0 to 250 Hz.

**TABLE-1**  
**Revised Details of Bid**

S. N.	Name of Equipment & Instruments	Bid Value (in Rupees)	Indicative Quantity (In no.)	Minimum average gross annual turnover for last three financial years (In Rupees)	Minimum average gross annual turnover for last three financial years for (MSMEs of Rajasthan) (In Rupees)	Bid Security (B.S) (In Rupees)	Bid Security (B.S) for MSME Units of Rajasthan (In Rupees)
1.	Single Channel ECG Machine	30000000	2300	18000000	9000000	600000	150000
2.	3 Channel ECG Machine	2600000	100	2500000	1250000	52000	13000
3.	12 Channel ECG Machine	6000000	100	2500000	1250000	120000	30000

Please note that all clarifications/amendments/corrigendum in technical specifications/bid conditions/bid details/date extension of bid are integral part of the bid document. This corrigendum/ addendum should be signed and annexed with bid document.

All other terms & conditions remain the same.

*Y M 28.1.2020*  
Executive Director (EPM)  
RMSCL, Jaipur

2/2