



Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Rajasthan, Jaipur

Phone No.-2223887, Fax no.-0141-2228065

E-Mail : edepmrmisc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website www.rmssc.health.rajasthan.gov.in

No. F-8(348)RMSC/EPM/M-2/2024-25/NIB-852

446

Date:- 28/11/24

Clarification/Corrigendum/Addendum

Subject:- NIB No. F-8(348)RMSC/EPM/M-2/2024-25/NIB-852/368 Dated: 25.09.2024 for item Incubating Flexible Laryngoscope/Fiber Optic Laryngoscope (Video Laryngoscope)

In Reference to subject cited above and NIB-852 The following Corrigendum/Addendum is issued for inclusion in bid document as below:-


1. REVISED TECHNICAL SPECIFICATION of "Video Laryngoscope"

1. It should be able to use on Adult&Pediatrics ..
2. It should have a portable full color TFT/LCD or better Display with minimum 60° field of view. Color display (TFT/LCD or better) attached to the handle; the dimension should be between minimum 2.4 inches (Diagonal) to maximum 3.5 inches (Diagonal).
3. Should have CMOS full color camera with integrated bright white LED Light source. For better handgrip there should not be any external wire/connection on handle grip.
4. It should have a Battery Status Indicator.
5. It should have Power saving mechanism to minimize battery consumption.
6. It should have automatic White Balancing (gain and exposure control) feature
7. It should be supplied with external display (TFT/LCD or better) of equal to or more than 7.5 inch (diagonally) with stand which can be connected through connecting cable or wireless. If required, Connecting Video outputs cable of at least 2 meters should be supplied .
8. The reusable display should come packaged in a protective case.
9. Should have provision to insert all sizes of endotracheal tube.
10. **It should have Rechargeable Lithium polymer Battery with 60 minutes operational time, charging adapter to charge the device should be supplied free of cost.**
11. It should be USFDA certified / European CE certified (from 04 digit notified body)/ BIS certified as per applicable Medical Device Class, related documents/certification shall be submitted along with technical bid.
12. Manufacturer/Importer should have quoted model Valid CDSCO registration (If applicable). Related documents should be submitted at the time of technical bid/during demonstration.
13. Metal Blade Types: Should have a robust anti-fogging property.
Should be supplied
 - a. Macintosh blade 1 size – 2 Nos.
 - b. Macintosh blade 2 Size -2 Nos.
 - c. Macintosh blade 3 Size -2 Nos.
 - d. Macintosh blade 4 Size - 2 Nos.
 - e. Macintosh D blade 4 Size -2 Nos.
14. Warranty: Warranty of Two years on complete system from the date of installation.
15. BOQ- Rate of extra blades to be quoted
16. BOQ- Rate of video stilllet compatible with the same screen to be quoted

2. Date extension of Bid:-

Existing Dates			Extended Dates		
Last Date & Time of Online Downloading of Bidding Document	Last Date & Time of Online Submission of Bid	Date & Time of Online Opening of Technical Bid	Last Date & Time of Online Downloading of Bidding Document	Last Date & Time of Online Submission of Bid	Date & Time of Online Opening of Technical Bid
28.10.2024 6:00 PM	28.10.2024 6:00 PM	29.10.2024 03:00 PM	06.11.2024 6:00 PM	06.11.2024 6:00 PM	07.11.2024 03:00 PM

All other terms & conditions remains the same.


Executive Director (EPM)
RMSCL, Jaipur

Note:- Please note that all above amendments/corrigendum is technical specifications/bid conditions is the integral part of (Section-VIII) and the bid document. This corrigendum/addendum should be signed and annexed with bid document. All other terms & conditions remains the same.