



Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Rajasthan, Jaipur
Phone No.-2223887, Fax no.-0141-2228065
E-Mail : edepmrmsc-rj@nic.in
CIN:U24232RJ2011SGC035067
Website www.rmssc.health.rajasthan.gov.in
No. F-8(350)RMSC/EPM/M-2/2024-25/NIB-855/447
Date:- 28/12/24

Clarification/Corrigendum/Addendum

Subject:- NIB No. F-8(350)RMSC/EPM/M-2/2024-25/NIB-855/381 Dated: 05.10.2024 for item (i) Essential Newborn Care and Resuscitation Mannequin & (ii) Breast Feeding Simulator

In Reference to subject cited above and NIB-855 The following Corrigendum/Addendum is issued for inclusion in bid document as below:-

1. Revised Specification:- GENERAL TECHNICAL SPECIFICATIONS OF ALLMANNEQUIN

1. The color of the mannequin should be simulating Indian babies/adult in medium skin tones.
2. The material of the mannequin should be of polyvinyl and silicon rubber free from any carcinogenic agents.
3. The texture of the mannequin should be soft and smooth and close to the feel of baby/adult skin as relevant. The texture must be friction free to demonstrate the desired procedure.
4. The internal parts of mannequin must be realistically sculpted, anatomically accurate and feel must be smooth/resilient/bony as relevant and suitable for simulation.
5. The mannequins must be portable and any fittings used in mannequins must be of aluminium or polycarbonate or equivalent. The mannequin's durability must be of minimum 3 years.
6. The material of the mannequin should withstand extremes of temperature (up to 45 degree Celsius).
7. The supplier must ensure manufacturer's guarantee against for the specification and also manufacturing defects.
8. All mannequins should include a soft/Hard Carrying Case and study questions, Dos and Don'ts. instructions manual, maintenance guide, background information, videotape for demonstrating the use of mannequin, User manual with troubleshooting guidance, technical manual with maintenance and first line technical intervention instructions and any other relevant teaching/training materials in English.
9. The mannequins should have additional accessories as listed and also talcum powder or silicone gel to avoid friction, list of accessories and spare parts with cost and contact details of its supplier preferably within State.

Technical Specification of Essential New Born care & Resuscitation Mannequin.

- a) The New born mannequin should be realistic in size and appearance and also natural weight, feel and touch.
- b) New born mannequin should have features for training essential new born care (ENBC) and new born resuscitation.

- c) New born mannequin should facilitate effective bag and mask ventilation, chest rise only with correct technique.
- d) The new born mannequin should include the following:
- I. Attached squeeze bulbs for simulation of cord pulsation, spontaneous breathing, auscultation of heart sound and cry.
 - II. 4 External umbilical cords and 6 umbilical ties
 - III. 4 baby sheets or towels.
 - IV. 2 Head cap.
 - V. 2 Neonatal mucus sucker (easy to open, clean, autoclavable and reusable)
 - VI. 2 Training Stethoscopes.
 - VII. Spare chest bag


TECHNICAL SPECIFICATIONS OF BREAST FEEDING SIMULATOR

- a) Breastfeeding Simulator should be a wearable simulator that allows highly realistic simulation of breast feeding technique, breast milk expression and kangaroo Mother Care.
- b) The materials of Breastfeeding Simulator should have mimic the adult female breast in term of shape and anatomical part. It should ideally be of neoprene fabric/Silicon, silicon nipples and facility to **shape** it on a person.
- c) The color of the simulator should ideally be skin tone nature.
- d) The simulator should have water pouches (with lock system to control flow of milk) which when filled give in the contours of breast and the water can be realistically expressed from the breast.
- e) The Breast Feeding Simulator should include the following:
- I. 30ml cup.
 - II. Transport Storage bag
 - III. 4 baby sheets or towels.
 - IV. Directions for Use.

2. Date extension of Bid:-

Existing Dates			Extended Dates		
Last Date & Time of Online Downloading of Bidding Document	Last Date & Time of Online Submission of Bid	Date & Time of Online Opening of Technical Bid	Last Date & Time of Online Downloading of Bidding Document	Last Date & Time of Online Submission of Bid	Date & Time of Online Opening of Technical Bid
28.10.2024 6:00 PM	28.10.2024 6:00 PM	29.10.2024 03:00 PM	06.11.2024 6:00 PM	06.11.2024 6:00 PM	07.11.2024 03:00 PM

All other terms & conditions remains the same.


 Executive Director (EPM)
 RMSCL, Jaipur

Note:- Please note that all above amendments/corrigendum is technical specifications/bid conditions is the integral part of (Section-VIII) and the bid document. This corrigendum/addendum should be signed and annexed with bid document. All other terms & conditions remains the same.