



Rajasthan Medical Services Corporation Limited, Jaipur
Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Rajasthan, Jaipur

Phone No.-2223887, Fax no.-0141-2228065

CIN:U24232RJ2011SGC035067

F-8(210) RMSC/EPM/M-2/NIB-491/2020-21/ 4726

E-Mail : edepmrmsc-ri@nic.in

Website: www.rmhc.health.rajasthan.gov.in

Dated: 08-09-2020

Clarification/Corrigendum/Addendum-II

Subject:-Revised Technical Specification for NIB No. F-8(210) RMSC/EPM/M-2/NIB-491/2019-20/3550 dated 17.02.2020 for item Haemoglobin Meter strip, Haemoglobin Meter Cuvette.

In Reference to subject cited above and NIB-491, earlier corrigendum/addendum was issued on dated 2.9.2020, in continuous of same, the following Corrigendum/Addendum is issued for inclusion in bid document as below:-

1. Revised Technical Specification of Haemoglobin Meter strip:-

Sr. No	Existing Technical Specification		Amended Technical Specification		
1.	It should work on single wavelength		It should work on single/dual wavelength		
2.	Sensitivity	More than 80%	Findings should be published in two peer reviewed indexed journals. The studies should be done in two different Indian settings by two independent teams of investigators.	Sensitivity	More than 80%
3.	Specificity	More than 80%		Specificity	More than 80%
4.	Bias (Limits of Agreement)	0.5 gm/dl(\pm gm/dl)		Bias (Limits of Agreement)	0.5 gm/dl(\pm gm/dl)

Please note that all clarification/amendment/corrigendum in technical specifications/bid conditions is the integral part of the bid document. This corrigendum/ addendum should be signed and annexed with bid document.

All other terms & conditions remains the same.

Executive Director (EPM)
RMSCL, Jaipur