



RAJASTHAN MEDICAL SERVICES CORPORATION LTD.
D- Block, SwasthyaBhawan, TilakMarg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN: U24232RJ2011SGC035067

E-Mail - edepmrmisc-rl@nlc.in

No. F-8 () RMSC/EPM/M-1/NIB-857/2024-25/488

Dated: 20/11/2024

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject:-Revised Technical Specification and Revised bid schedule of item (1) BIPAP Ventilator with well cushioned face mask and harness (2) Cushioned face mask with harness, under No. F-8 () /RMSC/EPM/M-1/NIB-857/2024-25/353 date: 19.09.2024

1. Revised Technical Specification:-

S.No.	Revised Technical Specification
1	Non-invasive ventilation support system to ensure proper ventilation to COPD patients.
2	It should be suitable for Adult and Pediatric patients.
3	Modes of operation: CPAP, Spontaneous, S/T, Time Mode & Pressure Control.
4	It should have display to monitoring facility for Tidal Volume, RR, Minute ventilation, I: E ratio.
5	Operating pressure range of IPAP : 4 to 30 cm H2O or more. EPAP : 4 to 20 cm H2O or more. CPAP : 4 to 20 cm H2O or more.
6	Unit should be capable of delivering Tidal Volume 1200 ml or more and have option to deliver guaranteed tidal volume.
7	Breath rate should be able to set up to 30 BPM or more with Ti of 0.5 to 3 sec. and rise time 100ms to 600ms.
8	Sensitivity settings: should have automatic trigger, cycle setting and leak compensation.
9	Unit should have integrated alarm for power, Pressure (High/Low), Low Minute ventilation, Apnea, Circuit disconnected etc. With digital technology for improved leak compensation during inspiration & exhalation.
10	Complete unit should be supplied in good quality bag which can carry adult and pediatric circuits with Adult and Pediatric masks along with equipment.
11	Quoted Model should have USFDA approval / CE certificate issued by a notified body registered in the European commission, under relevant category/BIS. Related documents / certification shall be submitted along with technical bid.
12	Quoted Model should have Valid CDSO Certified related documents shall be submitted along with technical bid.
13	Unit should have battery backup capacity of 2 hour or more.
14	Unit should be supplied with reusable adult and pediatric patient circuits and reusable full face mask with harness Adult & Pediatric (1 no. each), operational / training manual (1 no.).
15	Guarantee: Three years from the date of installation.
16	CMC: CMC shall be given @ 5% of net rate (excluding GST as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after Guarantee period of three years.

Signature valid

Digitally signed by Dr. Akash Alha
Designation: Executive Director
Date: 2024.11.20 17:36:49 IST
Reason: Approved





17	Installation will be done by supplier free of cost.
18	Firm should have service engineer network in Rajasthan.
19	The firm should submit technical compliance sheet along with catalogue as per technical specifications mentioning the make & model of quoted item in the Technical bid.

2. Revised Bid Schdeule—

E-bids are invited as per following revised time schedule:-

Existing Dates				Extended Dates			
Last date & time for sale of bid form	Last date & time of receipt of bid form	Last date of submission of bid document fees	Date & time of opening of technical bid	Last date & time for sale of bid form	Last date & time of receipt of bid form	Last date of submission of bid document fees	Date & time of opening of technical bid
<u>20.11.2024</u> <u>11:00 a.m.</u>	<u>20.11.2024</u> <u>6:00 p.m.</u>	<u>21.11.2024</u> <u>02:00 p.m.</u>	<u>21.11.2024</u> <u>03:00 p.m.</u>	<u>02.12.2024</u> <u>11:00 a.m.</u>	<u>02.12.2024</u> <u>6:00 p.m.</u>	<u>03.12.2024</u> <u>02:00 p.m.</u>	<u>03.12.2024</u> <u>03:00 p.m.</u>

Note:- Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-VIII) and the bid document. This corrigendum/addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

Executive Director (EPM)
RMSCL, Jaipur

Signature valid

Digitally signed by Dr. Akash Alha
Designation: Executive Director
Date: 2024.11.20 17:36:49 IST
Reason: Approved