



Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Rajasthan, Jaipur
Phone No.-2223887, Fax no.-0141-2228065
E-Mail : edepmrmisc-ri@nic.in
CIN:U24232RJ2011SGC035067 Website www.rmhc.health.rajasthan.gov.in
No.F-8(347) RMSC/EPM/M-2/NIB-847/2024-25/542 Date:- 12/12/24

Clarification/Corrigendum/Addendum

Subject:- NIB No.F-8(347) RMSC/EPM/M-2/NIB-847/2024-25/366 Dated: 24.09.2024
for item BP APPARATUS PROFESSIONAL

In Reference to subject cited above and NIB-847 The following Corrigendum/Addendum is issued for inclusion in bid document as below:-

1. REVISED TECHNICAL SPECIFICATION OF BP APPARATUS PROFESSIONAL

1. Model(s) to be designed for professional use in hospital settings (Personal home-based models will not be considered).
2. Quoted model meets standards of AAMI/ANSI/ISO 81060-2:2013 or as amended /BHS/ESH-IP (Valid document/certificate must be submitted in technical bid), device must have the validation as per international standards and the publication of the device validation should be available.
3. Pressure measurement range should be 0 to 280 mm Hg.
4. Pressure measurement accuracy of ± 5 mm Hg.
5. Pulse rate to be displayed.
6. Should have easily readable LCD/TFT/LED broad Display of 80mm or more Diagonally.
7. Measurement Method: Oscillometric Measurement.
8. Cuff Size: Cuff Sizes: Adult Normal (2 Nos), Adult Large(1 Nos) & Pediatric Small (1 Nos) to be supplied free of cost with each machine.
9. Device should have internally rechargeable battery of More than 150 measurements.
10. Should have AC Adaptor to be supplied free of cost with each machine
11. Detachable cuff/sleeve.
12. Built-in surge protection to prevent damage to instrument in case of power surge.
13. Three years warranty including all spares, accessories, consumables and re-calibration (Bidder has to replace the unit if not repairable within warranty period).
14. Devices should function between 5°C to 50 °C temperature.
15. Power Saving: Automatic power off if system idle for 3 to 5 minutes..
16. Minimum- Three years warranty including all spares, Accessories, consumables and recalibration and breakdown complaint must be attended within a week.
17. Quoted model should have US FDA/European CE with 4 digit notified body/BIS (Valid Certificate must be submitted in technical bid)
18. Device should have Bluetooth connectivity (optional).
19. Quoted model should have Valid CDSCO License, Copy must be submitted in technical bid.
20. Legal Metrology Licence for quoted model must be submitted in Technical Bid.
21. Low Battery indicator and error indicators should be available
22. Carrying case/bag to be provided free of cost with device.
23. Device should be handy & portable.
24. Services should be available pan Rajasthan till district level
25. No calibration needed up to 30,000 readings.
26. Device should have Data storage and transfer facility
27. Device should have Arrhythmia detection facility (Optional).

Signature valid

Digitally signed by Dr. Akash Alha
Designation: Executive Director
Date: 2024.12.12 17:42:03 IST
Reason: Approved

1. Date extension of Bid:-

Existing Dates			Extended Dates		
Last Date & Time of Online Downloading of Bidding Document	Last Date & Time of Online Submission of Bid	Date & Time of Online Opening of Technical Bid	Last Date & Time of Online Downloading of Bidding Document	Last Date & Time of Online Submission of Bid	Date & Time of Online Opening of Technical Bid
16.12.2024 6:00 PM	16.12.2024 6:00 PM	17.12.2024 03:00 PM	19.12.2024 6:00 PM	19.12.2024 6:00 PM	20.12.2024 03:00 PM

All other terms & conditions remains the same.

Executive Director (EPM)
RMSCL, Jaipur

Note:- Please note that all above amendments/corrigendum is technical specifications/bid conditions is the integral part of (Section-VIII) and the bid document. This corrigendum/addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

Signature valid

Digitally signed by Dr. Akash Alha
Designation: Executive Director
Date: 2024.12.12 17:42:03 IST
Reason: Approved