

## Rajasthan Medical Services Corporation Ltd.

D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)

Ph. No. 0141-2223887, Fax No. 0141-2228065 CIN:U24232RJ2011SGC035067 E-Mail - [rmisc@nic.in](mailto:rmisc@nic.in), [edepmrmisc-rj@nic.in](mailto:edepmrmisc-rj@nic.in)

No. F-8(195) RMSC/EPM/M-2/2020-21/NIB-470/ 7546

Date: 25/6/21

### Office Order


List of bidders declared responsive/non responsive as per documentary & technical evaluation in bid number **F-8(195) RMSC/EPM/M-2/2020-21/NIB-470/4649** Dated: **25.08.2020** after opening of technical bid on 25.11.2020 and as per decision taken in the purchase committee dated 28.05.2021 headed by M.D., RMSCL is under.

S. No.	Name of Firm	Name of Goods	Status of Documentary and Technical Level	Responsive or Non responsive and Reasons
1.	2.	3.	4.	5.
1	M/s Esteem Industries Inc. 2EF, EPIP, Phase-I, Jharmajri, Baddi-174103 Email:- <a href="mailto:info@esteem.ind.in">info@esteem.ind.in</a> <a href="mailto:Esteem.lab@gmail.com">Esteem.lab@gmail.com</a>	1.Hot Air Oven 2. Incubator 3.Incubator (Bacteriological) Digital	a) Documentary not approved (ISI Certificate not submitted by firm ) for item Hot Air Oven Demo not taken b) Documentary approved but demo not given by firm for item Incubator technically not approved. c) Documentary approved but demo not given by firm for item Incubator (Bacteriological) Digital technically not approved.	Non-Responsive - Non-Responsive Non-Responsive
2	M/s Genist Technocracy Pvt. Ltd. Barwala Road Village Saidpura, Opp. Mitter Dharam Kanta, Derabassi Dis. Mohali (PB.) – 140507 Email:- <a href="mailto:sales.genisttechnocracy@gmail.com">sales.genisttechnocracy@gmail.com</a>	1. Incubator 2. Incubator (Bacteriological) Digital	a)No Shortcoming document submitted on e-proc portal.	Non-Responsive
3	M/s Meril Diagnostics Pvt. Ltd. 2 <sup>nd</sup> floor, D1-D3, Meril Park, Survey No. 135/2/B & 174/2, Muktanand Marg, Chala, Vapi – 396191, Gujarat Email:- <a href="mailto:prakash.teckchandani@merillife.com">prakash.teckchandani@merillife.com</a>	Semi Automatic Elisa Reader and Washer	b)Documentary approved and Technically not approved due to disqualified in demo for item Semi Automatic Elisa Reader and Washer 1. The instrument memory is not as per specification 2. Liquid sensor not present in waste bottle 3. Plate sensor not available.	Non Responsive
4	M/s Sudarshan Chemicals Sudarshan Building, Jhalawar Road, Kota-324007 Email:- <a href="mailto:sudarshanchemicals@gmail.com">sudarshanchemicals@gmail.com</a>	1. Hot Air Oven 2. Incubator 3. Incubator (Bacteriological) Digital	a) Documentary not approved (ISI Certificate not submitted by firm ) for item Hot Air Oven Demo not taken. b) Documentary approved and Technically not approved for Incubator (Bacteriological) Digital. (Air circulating fan not available rubber gasket not available at the door)	Non-Responsive Non-Responsive

25/6/21

S. No.	Name of Firm	Name of Goods	Status of Documentary and Technical Level	Responsive or Non responsive and Reasons
1.	2.	3.	4.	5.
			c) Documentary and Technically approved for item Incubator .	Responsive
5	M/s Remi Elektrotechnik Ltd. Remi House, 11 Cama Indl. Estate, Goregaon (E), Mumbai – 400063 Email:- <a href="mailto:bids@remilabworld.com">bids@remilabworld.com</a>	1. Incubator	a)Documentary and Technically approved for Incubator .	Responsive
6	M/s S.R. Enterprises B-18 & 19, Jai Jawan Colony Ist,\ Tonk Road, Jaipur - 302018 Email:- <a href="mailto:srindiaentp@yahoo.com">srindiaentp@yahoo.com</a>	1. Fully Automated 4 Plate Elisa Processor	a)Documentary and Technically approved for item Fully Automated 4 Plate Elisa Processor.	Responsive
7	M/s Transasia Bio-Medicals Ltd. Transasia House, 8, Chandivali Studio Road, Andheri (East), Mumbai – 400072 Email:- <a href="mailto:transasia@transasia.co.in">transasia@transasia.co.in</a>	1. Fully Automated 4 Plate Elisa Processor, 2.	a) Documentary and Technically approved for item Fully Automated 4 Plate Elisa Processor.	Responsive
		Semi Automatic Elisa Reader and Washer	b) Documentary and Technically approved for item Semi Automatic Elisa Reader and Washer	Responsive
8	M/s Tulip Diagnostics (P) Ltd. 84, Lal Singh Building, 2 <sup>nd</sup> Floor, 205-206, Near Aggarwal Sweets, Vasant Kunj Road, Manipalpur, New Delhi – 110037 Email:- <a href="mailto:chnadra.mudgal@tulipgroup.com">chnadra.mudgal@tulipgroup.com</a>	1. Fully Automated 4 Plate Elisa Processor	a)No shortcoming document submitted on e-proc portal.	Non Responsive
9	M/s Yorco Sales Pvt. Ltd., Yorco Corporate Tower, Plot No. 2, Gazipur, Patparganj, Delhi-110096 Email:- <a href="mailto:sales@yorco.com">sales@yorco.com</a>	1. Hot Air Oven 2. Incubator 3. Incubator (Bacteriological) Digital	a) Documentary not approved (ISI Certificate not submitted by firm ) for item, Hot Air Oven Demo not taken	Non Responsive
			b) Documentary and Technically not approved for item Incubator Disqualified in Demonstration. (Cooling not available)	Non Responsive
			c) Documentary and Technically approved for Incubator (Bacteriological) Digital.	Responsive

Online financial bids of responsive bidders for above item will be opened on dated 29.06.2021 at 3:00 P.M.

  
**Executive Director (EPM)**  
**Rajasthan Medical Services Corporation Ltd.**  
**Jaipur**