



Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Rajasthan, Jaipur
Phone No.-2223887, Fax no.-0141-2228065

E-Mail : edepmrmc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website www.rmhc.health.rajabthan.gov.in

No. F-8(317) RMSC/EPM/M-2/NIB-730/2023-24/ 838

Date:- 7/12/23

Clarification/Corrigendum/Addendum

Subject:- NIB No. F-8(317) RMSC/EPM/M-2/NIB-730/2023-24/429 Dated 28.08.2023 for item Labour Table, Patient Strecher Trolley, Bed Side Locker, MVA Kit, Medicine Assorter Tray.

In Reference to subject cited above and NIB-730 The following Corrigendum/Addendum is issued for inclusion in bid document as below:-

1. Revised Specification:-

AMENDMENT AMENDMENT/REVISED TECHNICAL SPECIFICATIONS OF PATIENT STRECHER TROLLEY

1. Trolley made of pipe frame M.S. Tube 16 Gauge 1" round for horizontal & 1.25" for vertical.
2. Standard length & width and height 210cm $\pm 2\%$ (L) X 56cm $\pm 2\%$ (W) X 65cm (H) $\pm 2\%$.
3. Should have four noiseless 15 cm diameter heavy duty castor wheels, 2 Nos. with locks for break movement.
4. Wheels should not be welded out the trolley as repair / replacement of wheels later on is possible.
5. The trolley should have facility to keep Oxygen cylinder.
6. Should have weight bearing capacity of around at 200 Kg (Patient + O2 cylinder).
7. Head-end of the stretcher top size around 2.25 - 2.5 feet should be adjustable to put at deferent angles i.e.15 to 75 degree.
8. Detachable top should have this adjustable head end.
9. Arrangement for I/V stand and for supporting O2 (oxygen) cylinder.
10. 1.5" mattress at top, density of foam should be ≥ 40 .
11. Pre-treated and epoxy powder coated finished.

General Terms & Conditions:

1. Guaranty: Guaranty of one year against any manufacturing defect from the date of supply / delivery / installation.
2. Packing shall be of good quality to avoid damage or Scratches during transit.
3. Color of item shall be confirmed by ordering authority at the time of supply order or as per sample.
4. The final technical approval of furniture items shall be after inspection of samples by technical committee at the time of technical bid Evaluation.
5. The RMSCL may do pre dispatch inspection at supplier store/ premises/ factory.
6. Pre-Dispatch inspection of each supply shall be done by the technical committee and post-dispatch inspection on random sampling basis will be done as per the norms issued by RMSCL.

**AMENDMENT/REVISED TECHNICAL SPECIFICATIONS OF BED SIDE
LOCKER**

1. Overall Dimension:
 Length : 400 mm (± 2% variation)
 Width : 800 mm (± 2% variation)
 Height : 810 mm (± 2% variation)
2. Construction: - The outside diameter of the legs of the locker should be 25 x 25 x 1.20 mm Sq. tube and 1.22 mm thick. Height of the legs from the bottom should be at least 6"
3. The cabinet body made from CRC steel sheet in 1.0 mm thick.
4. Top should be made from 1.2 mm thick stainless steel sheet and it should have back and side bent upward for acting as a guard railing (height of the sheet not less than 3 inches) and front side should be bent downward.
5. Cabinet: -
 Height : 580 mm
 Width : 800 mm
 Depth : 390 mm
 Tolerance: (± 2% variation)
 Cabinet having two set of separate boxes of same height approximate (properly welded/fix partition).
6. Door: - single/double door, door of the cabinet should be made from CRC Sheet of 1.25 mm thick press bent to required size. Door of the cabinet is offered with 3 ventilation louvers to provide air vents. The door should be press bend in such manner on top/side of door to facilitate opening/closing.
7. Door is pivoted to the cabinet with help of stainless-steel rod.
8. Pivoted rod acting as a hinge and allow the door to swing.
9. Shoe made from hard rubber. Nominal height of the shoe should be 30 mm.
10. Finishing: - All Steel sheet components shall be made from CRCA sheet and have to be treated for the removal of free rust on the surface.
11. Pre-treated and epoxy powder coated should be not less than 50 microns..

4. Date extension of Bid:-

Existing Dates			Extended Dates		
Last Date & Time of Online Downloading of Bidding Document	Last Date & Time of Online Submission of Bid	Date & Time of Online Opening of Technical Bid	Last Date & Time of Online Downloading of Bidding Document	Last Date & Time of Online Submission of Bid	Date & Time of Online Opening of Technical Bid
07.12.2023 6:00 PM	07.12.2023 6:00 PM	08.12.2023 03:00 PM	14.12.2023 6:00 PM	14.12.2023 6:00 PM	15.12.2023 03:00 PM

All other terms & conditions remains the same.


Executive Director (EPM)
RMSCL, Jaipur

Note:- Please note that all above amendments/corrigendum is technical specifications/bid conditions is the integral part of (Section-VIII) and the bid document. This corrigendum/addendum should be signed and annexed with bid document. All other terms & conditions remains the same.