

Rajasthan Medical Services Corporation Limited, Jaipur
Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065
 CIN:U24232RJ2011SGC035067

E-mail : edepmrmsc-rj@nic.in
 Website:www.rmsc.nic.in

No F-8(28) RMSC/EPM/M-5/2016-17/NIB-186/ 4970

Dated: 23/11/16

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: Clarification related to Technical Specification of Bid Document for Equipping and Fabrication of ALS Ambulances under NIB-186 & corrigendum F-8 (28) RMSC/EPM/M-5/2016-17/NIB-186/4915 Dated 18.11.2016.

In reference to subject cited above & NIB-186, it is clarified that with regard to the technical specification in scope of fabrication work in ALS Ambulance, point no. 24 (fire extinguisher) is also added on page no. 05, in scope of fabrication work as follow.

24.Fire extinguisher


4- 5 kg capacity dry powder A B C type fire extinguisher to be fitted at LHS patient saloon extreme rear corner with the provision to hold by Velcro strips & a stainless steel holding bracket to be provided at floor has to be provided at floor. Fire extinguisher available with ambulance has to be suitably placed and fixed in driver compartment.

Note: - page no 10, point no. 2 replaced as follows,

Documents related to CE/European CE / USFDA / ISI etc (as applicable) for each equipment should be submitted in technical bid.

All other terms and conditions of Bid shall remain the same.

This bears the approval of M.D., RMSCL, Jaipur


Executive Director (EPM)
RMSCL, Jaipur