

## Rajasthan Medical Services Corporation Limited D-Block, Swasthya Bhawan, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax No. 0141-2228065

E-Mail - mdrmsc@nic.in; edepmrmsc-rj@nic.in

CIN: U24232RJ2011SGC035067

Website: www.rmsc.health.rajasthan.gov.in

F-8() RMSC/EPM/M-4/2020-21/NIB-510/ 83

Date: 09 10/2020

## Clarification/Corrigendum/Addendum

Subject: -Revised Technical Specification for F-8()RMSC/EPM/M-4/2020-21/NIB-510/4816 Dated 16.09.2020 for item Multiple Urine Strips (10 Parameters).

In Reference to subject cited above and NIB-510, the various representations received from the firms and issues raised by the Bidders are examined by the competent Authorities and technical committee. The following Corrigendum/Addendum is issued for inclusion in bid document as below:-

## Revised Technical Specification of Multiple Urine Strips (10 Parameters)

- 1. Urine strip use for semi quantitative analysis for 10 parameters:
  - (a) Glucose
- (b) Protein
- (c) pH
- (d) Specific
- (e) Nitrite
- Gravity /

- (f) Urobilinogen
- (g) Blood
- (h) Ketone
- (i) Leucocytes
- (j) Bilirubin
- 2. Interpretation of the result should be both by visual reading by comparing with color scale on label and strip analyser.
- 3. Strip should be CE certified and IVD approved.
- 4. It should read the result within 60 second.
- 5. It should have instruction manual (Hindi and English) for detail information regarding use of strip and principle of various parameters.
- 6. It should have following sensitivity /reference values:
  - a. Glucose

:<25mg/d1

b. Protein

:< 15 mg/dl

c. pH

: 5.5-7.0

d. Specific Gravity

: 1.015-1.025

e. Nitrite

:Negative :Positive

f. Urobilinogen

 $:< 17 \mu \text{ mol/L}$ 

g. Blood

: Negative : Positive {Or any other sensitive method(<5Ery/micro litre)}

h. Ketone

<2.0 mg/dl

- i. Leucocytes

- $< 10 leu/\mu l$
- Bilirubin
- : <0.2 mg/dle
- 7. Strip should be stable at  $+2^{\circ}$ C to  $+30^{\circ}$ C.
- 8. It should have pack size 1×100 strip.



## मुख्यमंत्री निःशुल्क जाँच योजन

- 9. Shelf life of the strip must be one year from the date of supply with open vial stability of strip must be at least three months.
- 10. Firm/Manufacturer should have its manufacturing & R&D unit in India or outside country and proof of the same have to be submitted.
- 11. Cost of one single strip should be provided in BOQ.
- 12. Demonstration of item is must for final approval.

Please note that all clarification/amendment/corrigendum in technical specifications/bid conditions is the integral part of the bid document. This corrigendum/ addendum should be signed and annexed with bid document.

All other terms & conditions remains the same.

Executive Director (EPM)
RMSCL, Jaipur