

मुख्यमंत्री निःशुल्क जाँच

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, SwasthyaBhawan, TilakMarg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

CIN:U24232RJ2011SGC035067

E-mail : edepmrmsc-ri@nic.in

Website: <http://rmsc.health.rajasthan.gov.in>

No. F-8(267) RMSC/EPM/M-2/NIB-470/2020-21/ 9108

Dated: 22/11/2021

All Principals Medical Colleges,
All the Superintendents of Medical Colleges & Attached Hospitals.
Director (PH/HA/IEC/AIDS/RCH)

Sub:- Approval/Purchase of Item, 'Elisa Reader', for the Rate Contract and supply, ending on 31.10.2023.

Please find enclosed herewith approved firm and **Elisa Reader** for the rate contract and supply ending on **30.11.2023**. The rate contract and its terms & conditions are also available on Rajasthan Medical Services Corporation Ltd., Website www.rmsc.health.rajasthan.gov.in

In reference to the above rate contract No. F-8(267) RMSC/EPM/M-2/NIB-470/2020-21/ 9108..... Dated 22/11/2021... the supplier has executed the agreement with RMSC, as enclosed herewith for your information & necessary action. This bears an approval of M.D., RMSCL, Jaipur.

Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd, Jaipur

No. F-8(267) RMSC/EPM/M-2/NIB-470/2020-21/ 9108

Dated: 22/11/2021

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS/Principal Secy/Secy. Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NHM Jaipur.
5. PA to Managing Director, RMSCL
6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Sciences, Jaipur.
10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zenana Hospital, MahilaChikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital- Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/ (RH)/(Hospital Admn.)
15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
16. All P.M.O.'s.
17. All CM&HO's.
18. All Dy. CM&HO's.
19. AGM (IT), RMSC, Jaipur
20. All Zonal BME are directed to carry out inspection of the supplied equipment as per RMSCL Letter no. 2513 dated 10-07-2018.
21. M/s Transasia Bio-Medicals Ltd. Transasia House, 8, Chandivali Studio Road, Andheri (East), Mumbai - 400072 Email:- transasia@transasia.co.in
22. Guard file.

Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd, Jaipur

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, SwasthyaBhawan, TilakMarg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

CIN:U24232RJ2011SGC035067

E-mail : edepmrmsc-rj@nic.in

Website: http://rmsc.health.rajasthan.gov.in

No. F-8(267) RMSC/EPM/M-2/NIB-470/2020-21/9109

Dated: 22/11/2021


Rate Contract and Price Details

S.N.	Item	Detail
1.	Name of Equipment / Item	Elisa Reader
2.	Name & Address of Firm	M/s Transasia Bio-Medicals Ltd. Transasia House, 8,Chandivali Studio Road, Andheri (East), Mumbai – 400072 Email:- transasia@transasia.co.in
3.	Bidder Status	Manufacturer (Non MSME)
4.	Performance Security	BG No. 5WSGT02212990006 Dated 26.10.2021 HDFC Bank Ltd., Mumbai, Amount in Rs. 67,000.00 (Rs. Sixty Seven Thousand Only)
5.	HSN Code	90278090
6.	Brand/ Make	M/s Transasia Bio-Medicals Ltd.
	Model	Erba/Lisa Scan EM
7.	Packing Unit	Each
8.	R/C No. & Validity	R.C. No. 721 (Valid Up to 30.11.2023)
9.	Indicative Quantity	30
10.	Rates for supply to RMSC from Suppliers	
	(i) Basic Rate (in Rs.)	187891.00
	(ii) IGST @ 18% (in Rs.)	33820.38
	(iii) CGST (in Rs.)	0.00
	(iv) SGST(in Rs.)	0.00
	(v) Approved Rate From Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	221711.38
11.	Rate For Supply to Consignee From RMSC	
	(i) Basic Rate (in Rs.)	187891.00
	(ii) RMSC Surcharge@5% On (i) (in Rs.)	9394.55
	(iii) CGST@ 9% { On (i)+(ii)} (in Rs.)	17755.70
	(iv) SGST@ 9% { On (i)+(ii)} (in Rs.)	17755.70
	(v) Approved Rate for Consignee from RMSC {(i)+(ii)+(iii)+(iv)} (in Rs.)	232796.95
12.	TECHNICAL SPECIFICATIONS OF ELISA READER	
	<ol style="list-style-type: none"> Halogen Lamp should have minimum usage option, i.e. it's should get 'on' while reading and should be 'off' mode even the instrument is on standby mode. Fully automatic & easy to operate. Photometer should be able to read all kind of wells flat, V and U. Photometer should be of 8 measurement channel. Should be able to read 2-3 strips or plate at a time. Should be totally open system. Should contain commonly required filter 405nm, 450nm, 630nm, (Space for optional filters Should have facility to measure at single as well as dual a wavelength (Bichromatic reading with photo metric accuracy of 2% single wavelength -15 seconds approx and 	

22/11/21

- wavelength -20 secs. approx
9. Should have built in thermal/matrix printer
 10. Should be able to report raw data, absorbance values, cut off and concentrations & multi standard
 11. Should be able to plot graphs and curves
 12. Absorbance range 0 to 4.0.
 13. Should be able to plot all kinds of curves like linear, log point to point it. should have large LCD display & memory should have plate making feature.
 14. It should be operate on 220V-230V ,50Hz AC Supply and it Should be supplied with power cord, spare fuses, dust cover, instruction manual and warranty card
 15. Quoted model should be CE and IVD/USFDA Approved (Certificates must be enclosed in technical bid).
 16. Multi label measurement and kinetic measurement
 17. Licensed window based software
 18. Display: Operation through PC
 19. Company should have local service network.
 20. Firm should provide original brochure of Equipment model with satisfactory users certificate and list.
 21. Should have automatic calibration and self diagnostics of any possible error in optical alignment before each program reading.
 22. Internal software (standalone mode) should have all the basic application related to ELISA like quality control, qualitative classification, standard graph etc.
 23. The instrument should have a memory of 99 inbuilt protocols in stand-alone mode and 100 result memory .
 24. System should have USB port for external printer connectivity optionfor the data printout.
 25. Firm should supply one desk top computer (System should have i3 processor with latest operating system, HDD 1 TB, 4GB RAM, 17" screen) with UPS. (Guarantee card with bills have to be submit with items to claim guarantee of these items).
 26. Compatible UPS with Battery should be provided free of cost with the system.
 27. **Warranty:** Three years on equipment from the date of installation.
 28. CMC will be given @5% of net rate-inclusive of GST (as applicable) & yearly escalation of 5% on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion guarantee period..

- The agreement signed between RMSCL and the firm forms integral part of the rate contract.


Executive Director (EPM)
RMSC, Jaipur


No. F-8(267) RMSC/EPM/M-2/NIB-470/2020-21/ 9109

Dated: 22/11/2021

Copy forwarded to the following for information & necessary action:

1. P.S. to Hon'ble Health Minister, Medical & Health Deptt. Govt. of Rajasthan.
2. P.S. to ACS/Principal Secy/Secy. Medical Health & FW Deptt. Rajasthan, Jaipur.
3. PS to Pr. Secy/Secy, Medical Education Department, Rajasthan, Jaipur
4. P.S. to Mission Director NHM Jaipur.

5. PA to Managing Director, RMSCL
6. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
7. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
8. Principal R.U.H.S. College of Dental Science Jaipur.
9. Principal, RUHS College of Medical Sciences, Jaipur.
10. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
11. Superintendent, SMS Hospital, Zenana Hospital, Mahila Chikitsalya, Mental Hospital, J.K. Lon Hospital, T.B & Chest Hospital, Jaipur.
12. Superintendent Associated Group of Hospital-Ajmer, M.G. Hospital, Jodhpur, Ummaid Hospital-Jodhpur, PBM Hospital-Bikaner, MBS Hospital-Kota & A.G. Hospital-Udaipur.
13. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
14. Addl. Director, Medical & Health Services (PW)/(RH)/(Hospital Admn.)
15. Joint Director, Medical & Health Services, (Zone) Jaipur, Ajmer, Udaipur, Jodhpur, Bikaner, Kota & Bharatpur.
16. All P.M.O.'s.
17. All CM&HO's.
18. All Dy. CM&HO's.
19. AGM (IT), RMSC, Jaipur.
20. All Zonal BME are directed to carry out inspection of the supplied equipment as per RMSCL Letter no. 2513 dated 10-07-2018.
21. M/s Transasia Bio-Medicals Ltd. Transasia House, 8, Chandivali Studio Road, Andheri (East), Mumbai - 400072 Email:- transasia@transasia.co.in
22. Guard file.


Executive Director (EPM)
RMSC, Jaipur