

RAJASTHAN MEDICAL SERVICES CORPORATION LTD.

D- Block, Swasthya Bhawan, Tilak Marg, C-Scheme JAIPUR (Raj)



Ph. No. 0141-2223887, Fax No. 0141-2228065

CIN: U24232RJ2011SGC035067

E-Mail - edepmrmisc-ri@nic.in

No. F-8 (67) RMSC/EPM/M-1/R.C.16-17/Glucom./NIB-181/2016/ 4066

Dated: 26/9/16

CLARIFICATION/CORRIGENDUM/ADDENDUM

Subject: - Amended Technical Specifications of Bid document of item Glucometer, under NIB No. F-8 (67) RMSC/ EPM /M-1 /R.C.16-17 /Glucom./NIB-181/2016/3802 Dated: 31.08.16 & Clarification/Corrigendum/Addendum No. F-8 (67) RMSC/EPM/M-1/R.C.16-17/Glucom./NIB-181/2016/3981 Dated:19.09.16

In reference to subject cited above and NIB-181, the representations received from the various firms and issues raised by the Bidders, are examined by the competent Authorities and technical committee. The following Clarification/Corrigendum/Addendum is issued for inclusion in Bid document Amended Technical Specification of items as below:-


1. Amended Technical Specifications of Items GLUCOMETER as below:-**REVISED TECHNICAL SPECIFICATION OF GLUCOMETER**

1. It should be a hand held meter.
2. It should be easy to use, two button navigation system.
3. It should have 96 segment LCD display with symbols.
4. It should be based on Electrochemical principal.
5. It should give fast and accurate result within 5-8 seconds (blood application with test strip within the meter)
6. No Coding system (Automatic calibration).
7. It should have under dosing detection system when less blood applied on the test strip.
8. Sample size should be 1-2 uL.
9. It should have measuring range between 60-350 mg/dL of blood sugar.
10. It should work in temperature range +10°C and +42°C and up to 85% humidity.
11. It should have Lithium coin battery.
12. It should have up to 500 test memory with date & time.
13. It should be ISO-15197:2013 approved, related documents should be submitted in Technical Bid.
14. Firm should provide accessories like test strip (50 Tests), lancing device (01), lancet (100nos), carry case (01), user manual and 1 lithium coin battery shall be supplied free of cost with every Glucometer.
15. Final approval after technical demonstration.
16. Guarantee shall be 3 year on Glucometer and 06 months on Lancing Device and Battery.
17. During this guarantee period the firm should provide compatible battery free of cost to consignee in case of battery failure.
18. Firm should calibrate the instrument in every 3 month for up to 3 years of Guarantee period.
19. Price of test strip (1 pckt = 50nos), lancing device (01 number) and lancet (1 pckt = 50nos) should be filled in Price Bid (BOQ).

It is also clarified that information of award of contract shall be communicated to all participating bidders on the website www.rmssc.nic.in and sppp.raj.nic.in. Please note that individual bidder will not be intimated."

Note:- Please note that all above amendments/corrigendum in technical specifications/bid conditions is the integral part of (Section-V, Schedule of Supply, and Point no. 3) and the bid document. This corrigendum/ addendum should be signed and annexed with bid document. All other terms & conditions remains the same.

This bears the approval of M.D., RMSCL, Jaipur.


Executive Director (EPM)
RMSCL, Jaipur