



Rajasthan Medical Services Corporation Limited, Jaipur

'D'-block, swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur

Phone no. 0141-2223887, Fax no. 0141-2228065
CIN : U24232RJ2011SGC035067

E-mail: edepmrmc-rj@nic.in
Website : www.rmcs.nic.in

No. F-8(40) RMSC/EPM/M-5/16-17/NIB-224/ 2501

Dated: 9/6/17


OFFICE ORDER

List of bidders declared responsive/non responsive as per documentary evaluation & technical evaluation of No. F-8(40) RMSC/EPM/M-5/16-17/NIB-224/646 Dated 17.02.17 technical bid opening date 12.04.17. The list of bidders declared responsive/ non responsive as per decision taken in the purchase committee dated 08.06.17 headed by M.D., RMSCL as follows:-

S no	Name of Firm	Name of Item	Responsive/ Non-Responsive	Remarks
1.	M/s Seerance Techno India Private Limited, 223 Chetak Chamber. Near Dawa Bazar, Indore-452001	Constant Temp. Water Circulation System	Responsive	NIL
2.	Sudarshan Chemicals, Sudarshan Building, Jhalawar Road, Kota	Gerber Centrifuge (Digital)	Responsive	NIL
		Laboratory Stirrer	Responsive	
		Constant Temp. Water Circulation System	Responsive	

Encloser - Demo Report Enclosed.

Online financial bids of responsive bidders for above item will be opened on dated 13.06.2017 at 11:00 A.M.


Executive Director (EPM)
Rajasthan Medical Services Corporation,
Jaipur

Technical Evaluation of LABORATORY STIRRER(NIB - 224)

Demo date: 26/05/2017

SON-52
 06/06/2017
 00-5

Sr. No.	Specifications	M/s. Sudarshan chemical, Kota
	Model & Make	ASE-173 d
1	Should have Maintenance free Brushless DC motor ensures high torque.	Yes
2	Should be compact Table Top Model , equipment basically consists of body, chuck, stainless steel blade propeller and a U-type stand.	Yes
3	Should have reverse/forward facility.	Yes
4	Should have 20 Liters Max. stirring capacity .	Yes
5	Should have Speed range 50 – 2200 rpm	Yes
6	Should have Monitoring of set & actual speed .	Yes
7	Equipment should be CE Certified .	Yes
8	Guarantee: Two years on equipment from the date of installation on all Parts of equipment including metal fitting / motor / Glass parts etc. against manufacturing defects	Yes
9	CMC: CMC shall be given @ 5 % of net rate (inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) . The CMC may be awarded for five years (on yearly basis) after Guarantee period of two years.	Yes
	Remarks	Technically recommended

6.6.17

616/17
 Chief Food Analyst
 Rajasthan, JAIPUR

06/06/17

Technical Evaluation of GERBER CENTRIFUGE (DIGITAL) (NIB - 224)

Demo date: 26/05/2017

Sr. No.	Specifications	M/s. Sudarshan chemical, Kota
	Model & Make	ASE-180
1	Should be electrically operated with timer 0-60 Minutes, adjustable in steps of 5 minutes.	Yes
2	should have Testing capacity-12 tests,	Yes
3	should be supplied with lactometer jar vessel made of metal having top finish of square and without a spout, internal diameter of 32±2mm and depth 185±5mm.	Yes
4	should have Mechanical Break for stopping	Yes
5	rotor should have bushing of brass .	Yes
6	should have effective radius 240-300mm and revolution 1020-1140 per minute .	Yes
7	should be supplied with Electronic Colony Counter.	Yes
8	Equipment should be CE certified.	Yes
9	Guarantee: Two years on equipment, accessories and all peripherals / sub systems from the date of installation, which Includes 2 visits and unlimited breakdown calls by service/application support engineer during the Guarantee period.	Yes
10	CMC: CMC shall be given @ 5 % of net rate (inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) . The CMC may be awarded for five years (on yearly basis) after Guarantee period of two years.	Yes
Remarks		Technically recommended

Handwritten signature and date: 26/5/17

Handwritten signature and date: 26/5/17
Chief Food Analyst
 Rajasthan, JAIPUR

Handwritten signature and date: 26/5/17

Technical Evaluation of Constant Temp. Water Circulation System (NIB - 224)
Demo date: 26/05/2017

Sr. No.	Specifications	M/s. Sudarshan chemical, Kota	M/s. Seerance Techno Pvt. Ltd., Indore
	Model & Make	ASE - 159 b	SCTB- 2G
1	Double walled construction, capacity 4 racks.	Yes	Yes
2	Stainless Steel -304 fabricated inner chamber and lid.	Yes	Yes
3	Outer chamber constructed with powder coated Stainless Steel -304.	Yes	Yes
4	Glass window fixed on the front for both sides, to provide a clear view during the immersion test.	Yes	Yes
5	To prevent heat loss, there is high grade glass wool between the chambers to minimize the loss of heat.	Yes	Yes
6	Working temp. range from ambient to 100°C, which is controlled by digital temp, indicator cum controller with $\pm 0.1^\circ\text{C}$ of accuracy.	Yes	Yes
7	For continuous stirring and maintaining uniform heating throughout the chamber, there is speed stirrer motor 1/20 HP is fitted in the unit.	Yes	Yes
8	The heating mechanism comprises of set of heaters.	Yes	Yes
9	Operates at 230 Volts AC only.	Yes	Yes
10	Equipment should be CE Certified.	Yes	Yes
11	Guarantee: Two years on equipment from the date of installation on all Parts of equipment including metal fitting / motor / Glass parts etc. against manufacturing defects	Yes	Yes
12	CMC: CMC shall be given @ 5 % of net rate (inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) . The CMC may be awarded for five years (on yearly basis) after Guarantee period of two years.	Yes	Yes
Remarks		Technically recommended	Technically recommended

Handwritten signature and date: 26/05/17

Handwritten signature and date: 26/05/17
 Rajasthan, JAIPUR

Handwritten signature and date: 26/05/17