



Rajasthan Medical Services Corporation Limited, Jaipur

'D'-block, swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur

Phone no. 0141-2223887, Fax no. 0141-2228065

E-mail: edepmrmssc-rj@nic.in

CIN : U24232RJ2011SGC035067

Website : www.rmssc.nic.in

No. F-8(40) RMSC/EPM/M-5/16-17/NIB-225/ 2503

Dated: 9/6/17

OFFICE ORDER

List of bidders declared responsive/non responsive as per documentary evaluation & technical evaluation of No. 8(40) RMSC/EPM/M-5/16-17/NIB-225/ 647 Dated 17.02.17 technical bid opening date 25.04.17. The list of bidders declared responsive/ non responsive as per decision taken in the purchase committee dated 08.06.17 headed by M.D., RMSCCL as follows:-

S no	Name of Firm	Name of Item	Responsive/ Non-Responsive	Remarks
1.	M/s Seerance Techno India Private Limited, 223 Chetak Chamber. Near Dawa Bazar, Indore-452001	Heating Mantles - 500 mL , 2 Litre (each)	Responsive	NIL
		Soxhlet Extraction Unit	Responsive	NIL
		U V. Cabinet with Long & Short wave UV Lamp	Responsive	NIL
2.	Sudarshan Chemicals, Sudarshan Building, Jhalawar Road, Kota	Heating Mantles - 500 mL , 2 Litre (each)	Non - Responsive	In Demonstration Technically Not Approved
		Soxhlet Extraction Unit	Responsive	NIL
		U V. Cabinet with Long & Short wave UV Lamp	Responsive	NIL
		Quartz Water Distillation Unit	Responsive	NIL

Encloser - Demo Report Enclosed.

Online financial bids of responsive bidders for above item will be opened on dated 13.06.2017 at 11:00 A.M.

202
Executive Director (EPM)
Rajasthan Medical Services Corporation,
Jaipur

Technical Evaluation of Heating Mantles - 500 mL, 2 Litre (each) (NIB - 225)

Demo date: 26/05/2017

Sr. No.	Specifications	M/s. Sudarshan chemical, Kota	M/s. Seerance Techno Pvt. Ltd., Indore
	Model & Make	ASE- 209 a & d	SHTM-5
1	Should be durable, chemical resistant polypropylene housing/non corrosive aluminum housing.	Yes	Yes
2	Flexible heating element ensures minimum risk of flask breakage.	Yes	Yes
3	These controlled and grounded electro mantles feature a built-in power regulator and safety earth (ground) screen. Unique ventilation slots or without ventilation slots keep the exterior safe to the touch even while in use.	Yes	Yes
4	The replaceable coiled heating element is suspended within a thermal insulating cartridge to provide maximum heat transfer and support.	Yes	Yes
5	Ceramic fiber insulation keeps the chemical-resistant polypropylene case cool.	Yes	Yes
6	Built-in power controller has dial markings from 1 to 10 for repeatability.	Yes	Yes
7	Mantles should have indicator lights for "power on" and "heater on".	Not Comply Only Power On	Power On Heater On
8	All mantles are double fused and include a single support rod clamp that accepts a 1/2" rod.	Yes	Yes
9	The stainless steel grounding screen provides added protection against shock in the event of spillage or flask breakage.	Yes	Yes
10	Mantles include 2-meter power cord with India plug.	Yes	Yes
11	Should be supplied 2 No's extra heating element with each Heating Mantels & distillation unit of borosil make 500mL, 2L (each).	Yes	Yes
12	Equipment should be CE Certified.	Yes	Yes
13	Guarantee: Two years on equipment from the date of installation on all Parts of equipment including metal fitting / tubes / Glass parts etc. against manufacturing defects.	Yes	Yes
14	CMC: CMC shall be given @ 5 % of net rate (inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) . The CMC may be awarded for five years (on yearly basis) after Guarantee period of two years.	Yes	Yes
	Remarks	Technically not recommended	Technically recommended

(Handwritten signature)

23/05/2017
6/6/17

(Handwritten signature)
26/06/17

Chief Food Analyst
Rajasthan, JAIPUR

Technical Evaluation of SOXHLET EXTRACTION UNIT (NIB - 225)

Demo date: 26/05/2017

Sr. No.	Specifications	M/s. Sudarshan chemical, Kota	M/s. Seerance Techno Pvt. Ltd., Indore
	Model & Make	ASE-213 b	SSEU-6 AG
1	The complete Soxhlet Extraction Apparatus consists of solvent reservoir extraction assembly with 6 extraction heaters to suit 250 ml. flasks	Yes	Yes
2	Outer casing is made of S.S.- 304 Steel.	Yes	Yes
3	Gap between casing & heater is filled with high grade glass wool insulation.	Yes	Yes
4	Extraction Heaters made of Fiber glass hand knitted netting having Nichrome heating elements.	Yes	Yes
5	Heating Elements having individual Regulators & Pilot Lamps to indicate working of the heaters.	Yes	Yes
6	Control box is away but Junctioned with the heating console.	Yes	Yes
7	Complete with Two retort rods, Cross bar, Clamp holder & Boss heads.	Yes	Yes
8	Operates on 230 Volts AC Single phase 50 Hz.	Yes	Yes
9	Should be supplied with heating element and 250ml. extraction units of borocil, suitable connection pipe 10 Meter.	Yes	Yes
10	Equipment should be CE Certified.	Yes	Yes
11	Guarantee: Two years on equipment from the date of installation on all Parts of equipment including metal fitting / tubes / Glass parts etc. against manufacturing defects	Yes	Yes
12	CMC: CMC shall be given (@ 5 % of net rate (inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) . The CMC may be awarded for five years (on yearly basis) after Guarantee period of two years.	Yes	Yes
Remarks		Technically recommended	Technically recommended

Handwritten signature and date: 26/05/17

**Chief Food Analyst
Rajasthan, JAIPUR**

Handwritten signature and date: 26/05/17

Handwritten signature and date: 26/05/17

Technical Evaluation of Quartz water distillation unit (NIB - 225)

Demo date: 26/05/2017

Sr. No.	Specifications	Model & Make	M/s. Sudarshan chemical, Kota
		ASE- 144	
1	Should have capacity of 1 Liter / Hrs.		Yes
2	Should be single stage distillation unit.		Yes
3	Should have manually or automatically fed facility into the distiller unit's boiling chamber.		Yes
4	Unit should have Quartz distiller with Demountable boiler.		Yes
5	Should have Compact design made by SS 304 material with Dual safety cutout.		Yes
6	High purity, low conductivity, Pyrogen free distillate.		Yes
7	Should be supplied with good Quality RO system 10 Liter / Hrs to connect before Distillation unit.		Yes
8	Equipment should be CE Certified.		Yes
9	Guarantee: Two years on equipment from the date of installation on all Parts of equipment including metal fitting / tubes / Glass parts etc. against manufacturing defects.		Yes
10	CMC: CMC shall be given @ 5 % of net rate (inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable). The CMC may be awarded for five years (on yearly basis) after Guarantee period of two years.		Yes
Remarks			Technically recommended

Handwritten signature and date: 26/5/17

Handwritten signature and date: 26/5/17
Chief Food Analyst
Rajasthan, JAIPUR


Handwritten signature and date: 26/05/17

Technical Evaluation of U. V. CABINET with Long & Short wave UV Lamp (NIB - 225)

Demo date: 26/05/2017

Sr. No.	Specifications	Model & Make	M/s. Sudarshan chemical, Kota	M/s. Seerance Techno Pvt. Ltd., Indore
1	U. V. Cabinet with short wave (254nm) long wave (365nm) wave, natural daylight tube in complete 4 mm thickness Black Acrylic Body for viewing and detecting separated substances on T. L. C. Layers & outer body should be M.S. powder coated.		ASE - 218	UV-C
2	Equipment should be CE certified.		Yes	Yes
3	Guarantee: Two years on equipment from the date of installation on all Parts of equipment including lamps / optical parts etc.		Yes	Yes
4	CMC: CMC shall be given @ 5 % of net rate (inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) . The CMC may be awarded for five years (on yearly basis) after Guarantee period of two years.		Yes	Yes
Remarks			Technically recommended	Technically recommended

Handwritten notes:
 CMC
 1.9.6

Signature: 
 21/5/17
Chief Food Analyst
 Rajasthan, Jaipur

Signature: 
 06/06/17