Alphabetical List of Surgical & Sutures NIT-1/2013/04 Dated 15/04/2013 (R.C.-III)

S. No.	Cod e No.	Name of approved item (s) with specification	Packi ng Unit	Name of Approved Firm	Approved Rate exclusive CST /VAT	Rate contract validity up to	Applicable rate of performance guarantee & amount of performance guarantee deposited	R.C. No.
1	S-5	 Disposable Sterile Surgical Rubber Gloves, Size 6.5 inches Made of natural rubber Latex, powdered without tear, properly folded in a paper Free of holes with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 Powder should be non-allergenic Conforms to IS 13422 ISI marked / CE certified / FDA approved 		M/s MRK Healthcare Pvt. Ltd., B4/B5, Byculla Service Industries, D.K. Marg, Sussex Road , Byculla, Mumbai 400 027 Email: info @mrkhealthcare.co m	Rs. 9.08 (Nine and Paisa eight only) + 2% CST	21.10.2014	B.G. no. 0062513BG0000051 dated 08.10.2013 Amount Rs. 25,00,000/- Bank - Centeral Bank of India, Mumbai Validity 07.10.2015 Performance Gurantee Rate 5%	III
2	S-6	 Disposable Sterile Surgical Rubber Gloves, Size 7 inches Made of natural rubber Latex, powdered without tear, properly folded in a paper Free of holes with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 Powder should be non-allergenic Conforms to IS 13422 ISI marked / CE certified / FDA approved 		M/s MRK Healthcare Pvt. Ltd.,Mumbai	Rs. 9.18 (Nine and Paisa eighteen only) + 2% CST	21.10.2014		III

3 S-	Disposable Sterile Surgical Rubber Gloves,	Pair	M/s MRK	Rs. 9.27 (Nine	21.10.2014	III
	 Size 7.5 inches Made of natural rubber Latex, powdered without tear, properly folded in a paper 		Healthcare Pvt. Ltd.,Mumbai	and Paisa twenty seven only) + 2%		
	 Free of holes with Acceptable Quality Level (AQL) of 1.5 or less Tensile strength as per EN 455-2 Powder should be non-allergenic Conforms to IS 13422 ISI marked / CE certified / FDA approved 			CST		