

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmssc-ri@nic.in

CIN:U24232RJ2011SGC035067


Website:www.rmssc.nic.in

No. F-8(45) RMSC/EPM/M-5/16-17/NIB-211/1140

Dated: 24/3/17

Rate Contract and Price details

S. No.	Item	Detail
1.	Name of Equipment / Item	Digital Thermometer
2.	Name & Address of Firm	M/s Microgene Diagnostic Systems Pvt Ltd, 806, Vikram tower, Rajendra Place, New Delhi-110008 Email: contact@microgenediagnostic.com
3.	Brand/ Model / Make	AccuSure MT1027
4.	Packing Unit	One
5.	RC No. & Validity	R.C. No- 370 & Valid upto 31.03.2019)
6.	Net Rate Per packing Unit (Rs.)	Rs. 55.44 (In words Rs. Fifty Five and paisa forty four Only)
7.	Taxes VAT/ CST (Rs.)	Exempted or as applicable
8.	RMSC Surcharges @ 5%	Rs. 2.77
9.	Total Price (Rs.)	Rs. 58.21
10.	Guarantee Period	One Year
11.	CMC after Guarantee Period	-
12.	<u>Technical Specification of Digital Thermometer</u>	
	<ol style="list-style-type: none"> 1. Digital LCD Display/at least 3 1/2 digit, Celsius Fahrenheit scale, accurate measurement: +/- 0.1°C between 89.6°F-107.6°F/32°C-42°C should be safe to use, no glass, no mercury. 2. Should have liquid crystal easy to read. 3. Should be water proof for ease of cleaning. 4. Should be battery powered with low battery indicator. 5. It should with clear instructions for use or preventive maintenance. 6. Can be calibrated in both centigrade and Fahrenheit, but if only one option is available, then Fahrenheit is preferable. 7. Should have buzzer signal function. 8. Should takes 60-90 seconds to measure temperature. 9. Can be used in the armpit/axilla, orally and rectally. 10. Registration and Model approval certificate from Department of Legal Metrology (Weights and Measure, Ministry of Consumer Affairs Government of India) is mandatory 	


Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd.,
Jaipur