

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmisc-ri@nic.in

CIN:U24232RJ2011SGC035067

Website:www.rmssc.nic.in

No. F-8()RMSC/EPM/M-2/2015-16/NIB-165/ 4342

Dated: 17/10/16

Rate Contract and Price details

S.N.	Item	Detail
1.	Name of Equipment / Item	Cold Box (Large)
2.	Name & Address of Firm	M/s Nilkamal Limited, Nilkamal House 77/78 Road No. 13/14, MIDC, Andheri (East), Mumbai- 400 093 E-mail:- info@nilkamal.com, furniture@nilkamal.com, govt.tenders@nilkamal.com
3.	Brand/ Model / Make	Nilkamal RCB444L
4.	Packing Unit	One
5.	R/C No. & Validity	R.C. No. 319 (Valid up to 30.04.2018)
6.	Net Rate Per packing Unit (Rs.)	Rs 6,435.83 (In Words, Rs. Six Thousand Four Hundred Thirty Five Paise Eighty Three Only) + VAT Exempted
7.	VAT	Exempted
8.	RMSC Surcharges @ 5%	Rs. 321.80
9.	Total Price (Rs.)	Rs. 6757.63
10.	Guarantee Period	3 years
11.	CMC after Guarantee Period	3 years
12.	<u>Specifications of item Cold Box (Large)</u>	
	<p>1 Description of Function</p> <p>1.1 Cold Box is essential for safe transportation of sensitive Food Item from the place of manufacturing to the place of field storage and final carriage to the place of Testing Lab. Thus CFC – free cold boxes ensure a pre-determined safe temperature range (+2 to + 8 C) for a particular period known as cold product classifications made by WHO.</p> <p>2 Operational Requirements</p> <p>2.1 The Large cold box, long range should be able to keep the Food Sample safe while transporting to long distances for longer hours of time.</p> <p>3 Technical Specifications</p> <p>3.1 Storage Capacity 20-30 Liters</p> <p>3.2 Weight Fully Loaded should be less than 50Kgs</p> <p>3.3 Weight Empty (with empty ice pack) should less 25Kg</p> <p>3.4 External surface and internal lining material LLDPE-(Linear Density Polyethylene)</p> <p>3.5 Insulation material CFC-free Polyurethane</p> <p>3.6 Insulation thickness 100-120 mm</p> <p>3.7 Each Cold Box shall contain adequate ice packs of following specs : Volume of ice Pack 0.35 -0.4 liters. External dimensions 165*95*33mm+/-1mm</p> <p>3.8 Each Cold box shall contain shall contain ice packs of WHO specifications PIS Ref : E/IP .2. The specification are as give below: Water Content: 0.35 to 0.4 Liters. External Dimension: 165x95x33 mm+/- 1 mm. Robustness: The Ice Packs samples shall withstand a one meter drop on every face, edge and corner when in a frozen state (-10 deg C to -20 deg C). It will then successfully pass the leakage test after thawing. Leakage Test: No leaks when 40 Kg lateral force is applied. Features: Effective reinforcements to restrain walls against swelling. Removable cap for filling ; cap to have effective internal water seal to resist 40</p> <p>3.9 Should be provided with handles for two people to carry it to a large vehicle.</p>	

✓ 2016

4 System Configuration Accessories, spars and consumables

4.1 Large Cold Box-Large Range (including integrated lockable fittings)

5 Power Supply: None**6 Standards, Safety and training**

6.1 The equipment should conform to WHO Specification E004/CB.01.3 The equipment should preferably be listed on the product Information Sheet of WHO.

6.2 The system should be tested as per WHO Standard Test procedures as per Eoo4/CB01-VP3 Copy enclosed.

7 Documentation

7.1 Manufacturer's certification of compliance of test procedures as per WHO Standards Test procedures.

7.2 **Certificate of Testing should be submitted at the time of supply from NABL certified lab / ILAC/ STQC Labs.****Terms & Conditions**

- Guarantee: Three years** on equipment from the date of installation.
- CMC:** CMC will be given @5% (of net rate-inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) and yearly escalation of 5% on last year's CMC price. The CMC may be awarded for three years (on yearly basis) after completion of guarantee period of three years.
- Installation and Training will be done by the supplier free of cost.
- The service engineer should be based in Rajasthan.
- The company should mention the make and model name/number of the quoted equipment and submit the technical brochure of the quoted model in the technical compliance sheet as per technical specifications.

13:

Comprehensive Maintenance Contract Charges (per year/per machine)

Rates (Rs.) Per Unit	First Year	Second Year	Third Year	Fourth Year	Fifth Year
Comprehensive Maintenance Agreement (CMA)	CMC: CMC will be given @5% (of net rate-inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) and yearly escalation of 5% on last year's CMC price. The CMC may be awarded for three years (on yearly basis) after completion of guarantee period of three years.				


 Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd.,

Jaipur

