

TECHNICAL SPECIFICATION
ABDOMEN HYSTERECTOMY SET

S.N.	Item	Qty
1.	RAMPLEY SPONGE HOLDING FORCEPS 25 CM STRAIGHT, SERRATED	2
2.	ROBIN TOWEL FORCEPS BLUNT 13.5CM	6
3.	SCALPEL HANDLE NO.3	1
4.	SCALPEL HANDLE NO.4	1
5.	MAYO HARRINGTON GYN. SCISSORS 18 CM STRAIGHT	1
6.	METZENBAUM-NELSON SCISSORS BL/BL 18 CM CURVED	1
7.	METZENBAUM-NELSON SCISSORS BL/BL 20 CM CURVED	1
8.	STANDARD TISSUE FORCEPS 20 CM STRAIGHT, 1X2 TEETH	1
9.	ALUS TISSUE SEIZING FORCEPS 15 CM 4X5 TEETH	6
10.	ALUS TISSUE SEIZING FORCEPS 20 CM, 4X5 TEETH	4
11.	BABCOCK INTESTINAL TISSUE FORCEPS 16 CM	2
12.	CRILE-RANKIN ARTERY FORCEPS 15 CM STRAIGHT	2
13.	CRILE-RANKIN ARTERY FORCEPS 15 CM CURVED	4
14.	DEAVERHARRIG RETRACTOR 1" X 25CM	1
15.	DEAVERHARRIG RETRACTOR 1.5" X 30CM	1
16.	OCHSNER-KOCHER ARTERY FORCEPS 20 CM STRAIGHT	2
17.	OCHSNER-KOCHER ARTERY FORCEPS 20 CM CURVED	4
18.	ALDERCREUTZ DISSECTING FORCEPS 15CM	1
19.	GEMINI FINE 90 DEGREE FORCEPS 20 CM	1
20.	CROSS HYSTRECTOMY CLAMP WITH SUTURE PASSER 20 CM CURVED	2
21.	BRIDGE HYSTRECTOMY CLAMP 20 CM STRAIGHT	2
22.	KILNER NEEDLE HOLDER 20CM	1
23.	FRITSCH RETRACTOR 24 CM 75 MM	1
24.	ABDOMINAL SPATULA 33 CM / 40 MM MALLEABLE	1
25.	HEANEY MAYO NEEDLE HOLDER 20 CM	1
26.	MAYO-HEGAR NEEDLE HOLDER 20 CM	1
27.	SUM SUCTION TUBE ALL S.S.	1
28.	ALDERCREUTZ DISSECTING FORCEPS 20CM	1
29.	PEAN ROCHESTER HEMOSTAT FCPS CURVED 20CM	4
30.	KIDNEY BOWL LENGTH 10" WIDTH 4" DEPTH 2"	1
31.	SOLUTION BOWL 4"	1

General Terms & Conditions:

1.	All surgical instruments should be CE certified & should be manufactured by ISO 9001:2008 & ISO 13485:2003 (valid documentation should be submitted in technical bid)
2.	Manufacturer should have ISO 7153 certificate, (valid documentation should be submitted in technical bid)
3.	Surgical instruments made by steel grade AISI410 and 420, test report should be submitted.(documentation should be submitted in technical bid)
4.	Laser marking along with code should be on the instrument.

2/12

ch

5.	All non-consumable articles (except glass or imported articles) like instruments and others accessories etc. should bear " Rajasthan Sarkar " on each item without which the supply will not be entertained.
6.	The company should submit technical compliance sheet as per technical specifications mentioning the make & model of quoted item in the Technical bid .
7.	Guarantee: One year against any manufacturing defect.
8.	Final technical approval after sample demonstration.

ch


Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd.,
Jaipur