

TECHNICAL SPECIFICATION
VAGINAL HYSTERECTOMY SET

S. N.	ITEM	Qt. Y
1.	MAYO LAXER DISSECTING SCISSOR STRAIGHT 18CM	1
2.	LAXER DISSECTING SCISSOR BL/BL CVD 20CM	1
3.	DRESSING FORCEPS STANDARD STR 18CM	1
4.	TISSUE FORCEPS 1X2T STR 18CM	1
5.	RUSSIAN GRASPING FORCEPS STR 20CM	1
6.	PEAN HEMOSTATIC FORCEPS FINE STR 14CM	2
7.	PEAN -ROCHESTER HEMOSTAT FCPS STR 15CM	4
8.	PEAN ROCHESTER HEMOSTAT FCPS CURVED 20CM	4
9.	SIMS DUCK BILLED VAGINAL SPECULAM MEDIUM	1
10.	OCHSNER-KOCHER HEMOSTATIC FORCEPS 1X2T STR 20CM	2
11.	OCHSNER-KOCHER HEMOSTATIC FORCEPS 1X2T CVD 20CM	4
12.	ROBIN TOWEL FORCEPS BLUNT 13.5CM	6
13.	MAYA – HEGAR NEEDLE HOLDER 20CM	2
14.	MAYA – HEGAR NEEDLE HOLDER 15CM	1
15.	SIMON ZERNY RETRACTOR 7"	1
16.	BRAUN TENACULUM FORCEPS 25CM	1
17.	SCHROEDER TENACULUM FORCEPS 23 CM SIDE CURVED	1
18.	ALLIS INTESTINAL FORCEPS 5x6T 20CM	2
19.	ALLIS INTESTINAL FORCEPS 4x5T 20 CM	6
20.	BABCOCK INTESTINAL TISSUE HOLDING FORCEPS 16 CM	2
21.	RAKIN BLADDER SOUND	1
22.	FE- MAL CATHETER	1
23.	KINDLY BOWL LENGTH 10" WIDTH 4" DEPTH 2"	1
24.	SOLUTION BOWL 4"	1
25.	BARD PARKER HANDLE No. 3, 4 & 7 No	1E Ach
26.	SPONGE HOLDING FORCEPS – 10"	3
GENERAL TERMS & CONDITIONS:		
1.	All surgical instruments should be CE certified & should be manufactured by ISO 9001:2008 & ISO 13485:2003 (valid documentation should be submitted in technical bid)	
2.	Manufacturer should have ISO 7153 certificate, (valid documentation should be submitted in technical bid)	
3.	surgical instruments made by steel grade AISI 410 and 420,	

2/2

ch

	test report should be submitted.(documentation should be submitted in technical bid)	
4.	Laser marking along with code should be on the instrument.	
5.	All non-consumable articles (except glass or imported articles) like instruments and others accessories etc. should bear "Rajasthan Sarkar" on each item without which the supply will not be entertained.	
6.	The company should submit technical compliance sheet as per technical specifications mentioning the make & model of quoted item in the technical bid	
7.	Guarantee: one year against any manufacturing defect.	
8.	Final technical approval after sample demonstration.	

ch


Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd.,
-Jaipur