

TECHNICAL SPECIFICATIONS
VAGINAL SPECULUM

Vaginal Speculum, Cusco pattern solid blades and folding handles and chromium plated large, medium and small

General Terms & Conditions for above item:

1. All surgical instrument should be CE certified (valid documentation should be submitted in technical Bid)
2. All surgical instrument should be manufactured by ISO 9001:2008 & ISO 13485:2003 (valid documentation should be submitted in technical Bid)
3. Manufacturer should have ISO 7153 certificate. (valid documentation should be submitted in technical bid)
4. Stainless Steel grade AISI 410 and 420, test report should be submitted.
5. Laser marking along with code should be on the instrument.
6. Guarantee: One year against any manufacturing defect.
7. The company should mention the make & model of the quoted items and submit the technical compliance sheet in technical Bid
8. Final technical approval after sample demonstration.





Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd
Jaipur