

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmssc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website:www.rmssc.nic.in

No. F-8() RMSC/EPM/M-2/2015-16/NIB-171/ 4332

Dated: 17/10/16

Rate Contract and Price details

S.N.	Item	Detail
1.	Name of Equipment / Item	Nest Baby Warmer
2.	Name & Address of Firm	M/s Phoenix Medical Systems Private Limited, No. 32/4 Jawaharlal Nehru Lal Salai, Ekkattuthangal P.O. Box-3205 Chennai-600032, E- mail:sales@pmsind.com
3.	Brand/ Model / Make	PHOENIX/ Embrace Nest
4.	Packing Unit	One
5.	R/C No. & Validity	R.C. No. 322 (Valid up to 30.04.2018)
6.	Net Rate Per packing Unit (Rs.)	Rs 31,861.00 (In Words, Rs. Thirty One Thousand Eight Hundred Sixty One Only) + 2% concessional CST against C-form
7.	Taxes Concessional CST @ 2% (Rs.)	Rs. 637.22
8.	RMSC Surcharges @ 5%	Rs. 1624.92
9.	Total Price (Rs.)	Rs. 34123.14
10.	Guarantee Period	3 Year
11.	CMC after Guarantee Period	5 Year
12.	<p align="center"><u>Specifications of item Nest Baby Warmer</u></p> <p>1, An easy to use precession heating system preferably electronically controlled with indicators of overheating / under heating.</p> <p>2 Specially designed to support/reinforce the new-born during transport, which is bacteria resistant, skin safe, and easy to clean. (size should be between 520mm x 250mm to 600mm x 300mm).</p> <p>3. Baby should be comfortable and safe away from electricity, hard surfaces or projections.</p> <p>4. A pouch of phase change Material (PCM) which maintains a temperature of ~37 degree Celsius for at least 4 hours. Size should be between 380mm x 220mm to 400mm x 250mm. (PCM should be certified nontoxic and nonhazardous. Also should ensure safety even in case of accidental damage to product)</p> <p>5. Weight of all components should be less than 5kg. Weight of baby-carrying and warming components should be less than 2 kg to allow transport and usage.</p> <p>6. Operating Conditions: Temperature: 10 – 34deg C Humidity: 30 – 75% RH.</p> <p>7. Operating Voltage: 230V/50Hz.</p> <p><u>General Terms & Conditions:</u></p> <p>1. Equipment should be US FDA approved or CE certified from notified body..</p> <p>2. Guarantee: Three years on equipment from the date of installation.</p> <p>3. CMC: CMC shall be given @ 5 % of net rate (inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after Guarantee period of three years.</p> <p>4. The company should mention the make & model name/number of the quoted equipment and submit the technical brochure of the quoted model in the Technical bid along with compliance sheet as per technical specifications.</p>	

13.	Comprehensive Maintenance Contract Charges (per year/per machine)					
	Rates (Rs.) Per Unit	First Year	Second Year	Third Year	Fourth Year	Fifth Year
	Comprehensive Maintenance Agreement (CMA)	CMC: CMC shall be given @ 5 % of net rate (inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after Guarantee period of three years.				

2 ms

Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd.,
Jaipur

[Signature]