

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmisc-ri@nic.in

CIN:U24232RJ2011SGC035067

Website:www.rmssc.nic.in

No. F-8() RMSC/EPM/M-4/16-17/NIB-201/ 1601

Dated: 13/4/17

Rate Contract and Price details

S. No.	Item	Detail
1.	Name of Equipment / Item	MVA KIT (Syringe with accessories).
2.	Name & Address of Firm	M/s Global Medical Device, Kalasmruti Building, Ground Floor, S.NO. 46/2 Dattaprasad Housing Society, Kharadi Road, Chandannagar, Pune 14 . E-mail: globalmedicaldevices@gmail.com Mobile No. 9890014348 Tel. 020-41208304
3.	Brand/ Model / Make	WOMENCARE GLOBAL
4.	Packing Unit	Each
5.	R/C No. & Validity	(R.C. No 378 /Valid upto 30.04.2019)
6.	Net Rate Per packing Unit (Rs.)	1814.00 (In Rs One Thousand Eight Hundred Fourteen Only) + (Concessional CST against C-form if applicable in percentage @ 2 %
7.	Taxes VAT/ CST (Rs.)	Rs. 36.28
8.	RMSC Surcharges @ 5%	Rs. 92.51
9.	Total Price (Rs.)	Rs. 1942.80
10.	Guarantee Period	-----
11.	CMC after Guarantee Period	-----
12.	<u>Technical specifications of MVA KIT (Syringe with accessories)</u>	
	<ol style="list-style-type: none"> 1. Syringe Type: Double valve manual vacuum aspiration syringe 2. Material: Plastic 3. Capacity: 50-60 cc. 4. Pressure: 24-26 inches of Mercury 5. Plunger: Double arm plunger with "O" Ring Collars stop along with retaining clips. 6. Valve: Made of Silicon double pinch valve with a liner. 7. Cannula: Malleable plastic, 4, 5, 6, 7, 8, 9, 10 and 12 mm (total = 8) in size, in 5 different colours. 8. Compatible with the syringe without an adapter. 9. Additional three (3) "O" rings & Silicon oil for lubrication, 3 (three) extra collar stop to be supplied with each kit. 10. Kit to be supplied in transparent plastic box which is also autoclavable. 	

- 11. Sterilization: All parts of the syringe should be cold sterilizable with Gluteraldehydye & 0.5% Chlorine, should be also autoclavable at 121°C
- 12. Reusability: The kit should be reusable at least for 25 times under defined sterilization conditions. A certificate in this regard should be furnished from a Government approved lab.
- 13. Shelf life: Minimum 2 yrs. from the date of delivery.
- 14. Manufacturer should be ISO 13485 certified.

13. **Comprehensive Maintenance Contract Charges (per year/per machine)**

Rates (Rs.) Per Unit	First Year	Second Year	Third Year	Fourth Year	Fifth Year
Comprehensive Maintenance Agreement (CMA)	--	--	--	--	--


Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd.,
Jaipur