

**TECHNICAL SPECIFICATIONS**  
**ELECTRIC BREAST PUMP (PROFESSIONAL USE)/HOSPITAL USE**

1. Should be BPA Free.
2. Should have soft round Funnel properly Air tight seal.
3. Should be one hand system.
4. Should have parts can be easily dismantled to make sterilization easy & quick.
5. **should be supplied with 12 bottles with cap (compatible with Pump. )**

**Terms & Conditions:**

6. Manufacturer should have ISO 9001 / ISO 13485 certified Certificate.
7. Equipment should be European CE/USFDA.
8. Manufacturer should provide Certificate of calibration with All Pumps.
9. Should be supplied with following Spares Parts / accessories like :
  - a) Membranes/ diaphragms - 4 per unit
  - b) Valve(If any) - 4 per unit
  - c) Breast shield (different sizes) - 10 per unit
  - d) Silicone tubing - 10 per unit
  - e) Tubing adapter (If any)- 4 per unit
  - f) Locking ring (If any) - 4 per unit
  - g) Locking adapter (If any) - 4 per unit.
10. **Guarantee:** Two years on equipment from the date of installation.
11. **CMC:** CMC will be given @ 5 % (of net rate - inclusive of Excise Duty & exclusive of VAT / CST etc.) Plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for three years (on yearly basis) after completion of Guarantee period of two years.
12. Installation & Training will be done by supplier free of cost.
13. The service engineer should be based in Rajasthan.
14. The company should mention the make & model name / number of the quoted equipment and submit the technical brochure of the quoted model in the Technical bid along with compliance sheet as per technical specifications.

  
**Executive Director (EPM)**  
**Rajasthan Medical Services Corporation Ltd.,**  
**Jaipur**