

**TECHNICAL SPECIFICATIONS**  
**FOWLERS BED**

1. Frame work of strong & precise steel tubes mounted on 10cms to 12.5 cms. dia swiveling castors.
2. Having four section top of perforated CRCA sheet duly machine press bent.
3. Backrest and knee-rest section maneuvered by two separate screws.
4. Tubular head Provision for four saline locations.
5. Thickness of sheet 18G.
6. Tubular head of MS 16G & 1.25" dia.
7. **Size:** 2000 to 2100 mm (l) x 800 to 900 mm (w) x 600 to 700 mm (h).
8. **Finish:** Should be preheated and Epoxy powder coated.

*cm*

*Dr B*

**Executive Director (EPM)**  
**Rajasthan Medical Services Corporation Ltd,**  
**Jaipur**