

**Rajasthan Medical Services Corporation Limited, Jaipur**  
Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur – 302005

Ph. No. 0141-2223887, Fax : 0141-2228065  
CIN:U24232RJ2011SGC035067

E-mail : edepmrmisc-rj@nic.in  
Website: www. rmisc.health.rajasthan.gov.in

F-8(205)RMSC/EPM/M-3/16-17/NIB-207/

2317

Dated: 26-5-17

**Rate Contract and Price Details**

S.N.	Item	Detail
1.	Name of Equipment/Item	Table Post-Martum
2.	Name & Address of Firm	M/s United Surgical Industries, 445, F.I.E. Patparganj Industrial Area, Delhi-110092 (India) Email:-info@unisurg.com
3.	Brand/ Model / Make	United USI-2000
4.	Packing Unit	Each
5.	R/C No. & Validity	R.C. No. 389 (Valid up to 31.5.2019)
6.	Net Rate Per packing Unit (Rs.)	40810.00 (In words Rs. Forty Thousand Eight Hundred Ten Only)
7.	Taxes Concessional CST @ 2% (Rs.)	Rs. 816.20
8.	RMSC Surcharges @ 5%	Rs. 2081.31
9.	Total Price (Rs.)	Rs. 43707.51
10.	Guarantee Period	Guarantee: one year from the date of successful installation at consignee place. The firm shall install this item at free of cost.
11.	<b>Technical Specification of POST MORTUM TABLE</b>	
	<p>1. SIZE: Width :- 102", Depth:- 30", Height :- 34"</p> <p>2. TOP: 1.5mm thick S.S. secured to framework. All ends shall be closed, welded, ground and polished smooth. Top should be sound deadened.</p> <p>3. FRAMEWORK: 38 mm x 38 mm x 3 mm S.S. welded angles, secured to top and reinforced with 1.5mm thick S.S: 25 mm x 100 mm x 25 mm channels.</p> <p>4. SINK : One no. of 1.5mm thick Polished S.S. fully covered integral with work top, 16" x 16" x 8" Deep. Supplied complete with 50 mm dia waste and deck mounted faucet.</p> <p>5. LEGS: Six no. of legs and all legs to be constructed of not less than 40 mm x 40 mm square 1.5mm thick S.S. Stainless steel pipe. All legs shall be uniform in finish. All legs shall be provided with stainless steel bullet feet, having an integrally formed shaft, with a minimum adjustment of approximately 40-mm without using of threading or adjusting bolts. Feet shall be completely sealed at bottom.</p> <p>6. CROSSBRACING: All leg cross bracing, where required are to be constructed of not less than 25 x 25 mm square 1.2 S.S. Pipe. All cross bracing to run horizontal between all legs, approximately 150 mm above floor, unless otherwise specified. All joints to be completely welded around entire perimeter, forming complete seal, with all weld ground and painted.</p> <p>S. S. wherever used, shall be 304 grade. Relative documents shall be submitted in technical bid.</p> <p>8. Guarantee: one year from the date of successful installation at consignee place. The firm shall install this item at free of cost.</p>	

  
Executive Director (EPM)  
Rajasthan Medical Services Corporation Ltd.,  
Jaipur