


SPECIFICATION OF IV STAND

1. Stainless Steel (S.S.) Tubular pipe frame with S.S. nut.
2. Heavy duty plastic / S.S. base on five castors (50mm) for stability (Non rusting).
3. 2 hooks provided made up of SS rod for I.V. fluids.
4. Thickness of S.S. Tubular pipe 16 Gauge of 1.25" diameter.
5. S.S. adjustable rod from 135 to 240 cms.

General Terms & Conditions:

1. Guaranty: Guaranty of one year against any manufacturing defect from the date of supply / delivery / installation.
2. Packing shall be of good quality to avoid damage or scratches during transit.
3. Color of item shall be confirmed by ordering authority at the time of supply order or as per sample.
4. The final technical approval of furniture items shall be after inspection of samples by technical committee at the time of technical bid evaluation.
5. The RMSCL may do pre dispatch inspection at supplier store/premises/factory.



Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd.,
Jaipur

