

SPECIFICATION OF REVOLVING STOOL

1. M.S. tubular four legs (1.25" OD x 16 Gauge thick tube) fitted with PVC stumps.
2. Foot ring support of stainless steel.
3. 18" to 27" adjustable height S.S. top (18 gauge thickness) through screw system & it should be stable at max. height.
4. Pre-treated and epoxy powder coated.

General Terms & Conditions:

1. Guaranty: Guaranty of one year against any manufacturing defect from the date of supply / delivery / installation.
2. Packing shall be of good quality to avoid damage or scratches during transit.
3. Color of item shall be confirmed by ordering authority at the time of supply order or as per sample.
4. The final technical approval of furniture items shall be after inspection of samples by technical committee at the time of technical bid evaluation.
5. The RMSCL may do pre dispatch inspection at supplier store/ premises/ factory.



Executive Director (EPM)

**Rajasthan Medical Services Corporation Ltd.,
Jaipur**

