

**Rajasthan Medical Services Corporation Limited, Jaipur**

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

CIN:U24232RJ2011SGC035067

E-mail : edepmrmrsc-rj@nic.in

Website:www.rmrc.nic.in

No. F-8 (99) RMSC/EPM/M-1/R.C.2016-17/2016 / 4587

Dated: 27/10/18

**Rate Contract and Price Details**

S.N.	Item	Detail
1.	Name of Equipment / Item	Micro-plate Centrifuge
2.	Name & Address of Firm	M/s Vitan Medical Systems Pvt. Ltd. GF 2&4, Prabhat Kiran Building, Rajendra Place, New Delhi E-mail: info@vmsl.org Phone: 45094741 , 45094742 Fax:91-11-25724135
3.	Brand/ Model / Make	I D Centrifuge 12S II Bio -Red
4.	Packing Unit	Each
5.	R/C No. & Validity	R.C. No. 334 (Valid up to 30.04.2018)
6.	Net Rate Per packing Unit (Rs.)	Rs. 3,45,000.00
7.	Taxes Consessional CST @ 2%/VAT @ 5% (Rs.)	Rs. 6900.00
8.	RMSC Surcharges @ 5%	Rs. 17595.00
9.	Total Price (Rs.)	Rs. 369495.00
10.	Guarantee Period	Year :-Three year, from the date of installation of Equipment
11.	CMC after Guarantee Period	CMC :- CMC will be given @ 5 % (of net rate - inclusive of Excise Duty & exclusive of VAT / CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of three years
12.	<b><u>Specifications of Item microplate centrifuge (Immuno-hematologic gel micro column card centrifuge)</u></b>	
	<p><b>1. Purpose of Equipment:</b></p> <p>a) Immuno-hematologic Gel-microcolumn-Card-centrifuge to perform manual centrifugation step for Blood Grouping, Cross Matching, antibody screening or identification or phenotyping by coombs and enzyme phase by gel microcolumn technique to detect both IgG &amp; IgM antibodies, and also potentially usable for C3d, Partial/weak D, Single Rare antigens, PNH, Heparin/PF4 Ab Test (HIT), Syphilis antibody test etc.</p> <p>b) Must be designed specifically for blood bank use. Commercial or modified commercial centrifuges for other purpose are not acceptable</p> <p><b>2. Quality Standard:</b></p> <p>a) Manufacturing should be compliant with ISO 13485 (Valid documentation should be submitted in technical bid).</p> <p>b) Should be compliant with European CE with notified body number according to IVD Directive 98/79/EC or US FDA for this specific purpose. (Valid documentation should be submitted in technical bid).</p> <p>c) Equipment must be certified for electrical safety specifications of IEC/TR 60601-3-020: "Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 3-020: Conformity verification report for IEC 60601-2-020:1992 Particular requirements for laboratory centrifuges"</p> <p><b>3. Capacity, Construction and Functioning</b></p> <p>a) Centrifuge head should have minimum 12 slots to accommodate 12 of corresponding manufacturer's immuno-hematologic Gel microcolumn cards.</p> <p>b) Swing out suspensions for Gelcard slots</p> <p>c) Aerodynamic compact construction with vibration free performance; Noise level should be less than 60dB.</p> <p>d) Bottom of the microcolumn should have a conical (v) shape, u shaped bottom is not acceptable.</p> <p><b>4. Lid</b></p> <p>a) The lid of the centrifuge should be transparent and should have auto-locking during spinning</p> <p><b>5. Electrical characteristics:</b></p> <p>a) Must be compatible with Input voltage: 220/240V 50/60 Hz Ac</p> <p>b) Should have an integrated voltage stabilizer or should come with external stabilizer.</p>	

*(Handwritten signatures and initials)*

	<p>c) Microprocessor controlled programming with LCD screen displaying Rpm or RCF, time and other functions should be displayed real time</p> <p>6. <b>Guarantee:</b> Three years on equipment from the date of installation.</p> <p>7. <b>CMC:</b> CMC will be given @ 5 % (of net rate - inclusive of Excise Duty &amp; exclusive of VAT / CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of three years.</p> <p>8. The company should submit technical compliance sheet as per technical specifications mentioning the make &amp; model of quoted item along with catalogue in the Technical bid.</p> <p>9. Installation will be done by supplier free of cost.</p> <p>10. The service engineer should be based in Rajasthan.</p> <p>11. Four Preventive Maintenance services annually during Guarantee and CMC period are essential.</p> <p>12. Demonstration of equipment is must for final technical approval.</p> <p>13. <b>Consumable:</b> Consumable kit. Rate of Consumable kit should be provided by bidder in BOQ.</p>					
13.	<b>Comprehensive Maintenance Contract Charges (per year/per machine)</b>					
	<b>Rates (Rs.) Per Unit</b>	<b>First Year</b>	<b>Second Year</b>	<b>Third Year</b>	<b>Fourth Year</b>	<b>Fifth Year</b>
	<b>Comprehensive Maintenance Agreement (CMA)</b>	<p><b>CMC:</b> CMC will be given @ 5 % (of net rate - inclusive of Excise Duty &amp; exclusive of VAT / CST etc.) Plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of three years.</p>				

*Meer*

*22*

**Executive Director (EPM)**  
**Rajasthan Medical Services Corporation Ltd.,**  
**Jaipur**

*Chai's*