

**Rajasthan Medical Services Corporation Limited, Jaipur**

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmsc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website:www.rmssc.nic.in

No. F-8(87) RMSC/EPM/M-1/R.C.2016-17/2016 / 4569

Dated: 27/10/16

**Rate Contract and Price Details**

S.N.	Item	Detail
1.	Name of Equipment / Item	Plasma Expresser (Manual)
2.	Name & Address of Firm	M/s Terumo Penpol Private Limited P.B. No. 6105, I-2, Jawahar Nagar, Thiruvananthapuram, Kerala E-Mail:- <a href="mailto:info@terumopenpol.com">info@terumopenpol.com</a> Telephone Number:- 91 471 3015500 Fax:-91 471 2721519
3.	Brand/ Model / Make	M- 1.0 Terumo
4.	Packing Unit	Each
5.	R/C No. & Validity	R.C. No. 325 (Valid up to 30.04.2017)
6.	Net Rate Per packing Unit (Rs.)	Rs. 15,149.00
7.	Taxes Consessional CST @ 2% (Rs.)	Rs. 302.98
8.	RMSC Surcharges @ 5%	Rs. 772.60
9.	Total Price (Rs.)	Rs. 16224.58
10.	Guarantee Period	Year :- Two years, from the date of installation of Equipment
11.	CMC after Guarantee Period	CMC :- CMC will be given @ 5 % (of net rate - inclusive of Excise Duty & exclusive of VAT / CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of two years
12.	<p align="center"><b><u>Specifications of item Plasma Expresser (Manual)</u></b></p> <ol style="list-style-type: none"> <li>Should be suitable to extract blood components (plasma, platelets) from collection container</li> <li>Mode of operation: Manual</li> <li>Front panel should be spring loaded to uniform pressure on container causing transfer of fluid</li> <li>Compression plate should be made of transparent acrylic, durable</li> <li>Metal used for the equipment should be non-corrosive and can be cleaned with antiseptics.</li> <li>Base portion and Vertical surface should be made to have better strength and long lasting performance.</li> <li>The system should be CE approved (Valid documentation should be submitted in technical bid).</li> <li>Manufacturing should be compliant with ISO 13485 (Valid documentation should be submitted in technical bid).</li> <li><b>Guarantee:</b> Two years on equipment from the date of installation.</li> <li><b>CMC:</b> CMC will be given @ 5 % (of net rate - inclusive of Excise Duty &amp; exclusive of VAT / CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of two years.</li> <li>The company should submit technical compliance sheet as per technical specifications mentioning the make &amp; model of quoted item along with catalogue in the Technical bid</li> </ol>	

*[Signature]*

*[Signature]* 2/2

- 12. Installation will be done by supplier free of cost.
- 13. The service engineer should be based in Rajasthan.
- 14. Two Preventive Maintenance services annually during Guarantee and CMC period are essential.
- 15. Demonstration of equipment is must for final technical approval

13. **Comprehensive Maintenance Contract Charges (per year/per machine)**

Rates (Rs.)Per Unit	First Year	Second Year	Third Year	Fourth Year	Fifth Year
<b>Comprehensive Maintenance Agreement (CMA)</b>	<b>CMC:</b> CMC will be given @ 5 % (of net rate - inclusive of Excise Duty & exclusive of VAT / CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of two years				

*Manoj*

*[Signature]*

**Executive Director (EPM)**  
**Rajasthan Medical Services Corporation Ltd.,**  
**Jaipur**

*[Signature]*