

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmisc-rj@nic.in

CIN:U24232RJ2011SGC035067

Website:www.rmssc.nic.in

No. F-8(97) RMSC/EPM/M-1/R.C.2016-17/2016 / 4589

Dated: 27/10/16

Rate Contract and Price Details

S.N.	Item	Detail
1.	Name of Equipment / Item	Double Pan Balance
2.	Name & Address of Firm	M/s Remi Elektrotechnik Ltd., Remi House, 11 CAMA INDL Estate, Goregaon (E), Mumbai-400063 Ph. 91-22-4058 9888 Fax: 91-22-4058 9890 E-mail : sales@remilabworld.com
3.	Brand/ Model / Make	CBZ-50 Plus, Remi
4.	Packing Unit	Each
5.	R/C No. & Validity	R.C. No. 335 (Valid up to 30.04.2018)
6.	Net Rate Per packing Unit (Rs.)	Rs. 47,352.38
7.	Taxes Consessional CST @ 2%/	Rs. 947.05
8.	RMSC Surcharges @ 5%	Rs. 2414.97
9.	Total Price (Rs.)	Rs. 50714.40
10.	Guarantee Period	Year :-Two year, from the date of installation of Equipment
11.	CMC after Guarantee Period	CMC :- CMC will be given @ 5 % (of net rate - inclusive of Excise Duty & exclusive of VAT / CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of two years
12.	<u>Technical Specifications of Item Double PAN Balance</u>	
	<ol style="list-style-type: none"> 1. 2 independent weight sensors & display weight differences range 0 to 2500 gms 2. Weight measurement range 0 to 2500 gms 3. Weight sensors 4. Accuracy: ± 2 gms 5. LED/LCD Display with facility after Equal balance and Pan wt. difference 6. Automatic "0" facility /Auto Calibration facility 7. Accessories like power adaptor 8. Accuracy - Wt. Calibration tolerance : ± 1gm - Total Accuracy of display : ± 0.1 gm Power source-220-240 Ac 50Hz 9. Should have provision to convert weight to volume as per the specific gravity of all the blood components for the both Pan. 10. Double Pan Balance should have transparent lid cover for protection when not in use. 11. The system should be CE approved (Valid documentation should be submitted in technical bid). 12. Manufacturing should be compliant with ISO 13485 (Valid documentation should be submitted in technical bid). 13. Guarantee: Two years on equipment from the date of installation. 14. CMC: CMC will be given @ 5 % (of net rate - inclusive of Excise Duty & exclusive of VAT / CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of two years. 15. The company should submit technical compliance sheet as per technical specifications mentioning the make & model of quoted item along with catalogue in the Technical bid 16. Installation will be done by supplier free of cost. 17. The service engineer should be based in Rajasthan. 18. Two Preventive Maintenance services annually during Guarantee and CMC period are essential. 19. Demonstration of equipment is must for final technical approval 	

[Signature]

[Signature] 2/12

13. Comprehensive Maintenance Contract Charges (per year/per machine)

Rates (Rs.) Per Unit	First Year	Second Year	Third Year	Fourth Year	Fifth Year
Comprehensive Maintenance Agreement (CMA)	CMC: CMC will be given @ 5 % (of net rate - inclusive of Excise Duty & exclusive of VAT / CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of two years.				

Handwritten mark

Handwritten signature

Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd.,
Jaipur

Handwritten signature