

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmrsc-ri@nic.in

CIN:U24232RJ2011SGC035067

Website:www.rmssc.nic.in

No. F-8(92) RMSC/EPM/M-1/R.C.2016-17/2016/4579

Dated: 27/10/16

Rate Contract and Price Details

S.N.	Item	Detail
1.	Name of Equipment / Item	Sterile Connecting Device
2.	Name & Address of Firm	M/s Terumo Penpol Private Limited P.B. No. 6105, I-2, Jawahar Nagar, Thiruvananthapuram, Kerala E-Mail:- info@terumopenpol.com Telephone Number:- 91 471 3015500 Fax:-91 471 2721519
3.	Brand/ Model / Make	TSCD -II Terumo
4.	Packing Unit	Each
5.	R/C No. & Validity	R.C. No. 329 (Valid up to 30.04.2018)
6.	Net Rate Per packing Unit (Rs.)	Rs. 5,80,000.00
7.	Taxes Consessional CST @ 2% (Rs.)	Rs. 11600.00
8.	RMSC Surcharges @ 5%	Rs. 29580.00
9.	Total Price (Rs.)	Rs. 621180.00
10.	Guarantee Period	Year :- Three years, from the date of installation of Equipment
11.	CMC after Guarantee Period	CMC :- CMC will be given @ 5 % (of net rate - inclusive of Excise Duty & exclusive of VAT / CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of three years
12.	<u>Specifications of item Sterile Connecting Device</u>	
	<p>1. Purpose of Equipment</p> <p>i. Sterile Connecting Device is able to heat and cut two tubes and switch the connectivity of the two halves and weld them back, and do the cutting and welding in such a manner that it can be considered as sterile as fully closed process</p> <p>2. Quality Standard</p> <p>1. Manufacturing should be compliant with ISO 13485 (Valid documentation should be submitted in technical bid).</p> <p>2. Should be compliant with European CE with notified body number or US FDA (Valid documentation should be submitted in technical bid).</p> <p>3. Equipment must meet electrical safety specifications of IEC 60601-1</p> <p>3. Operational requirements</p> <p>a) Should ensure sterile connection between tubing to enable transfer of fluid/ blood from one container/ bag to another by welding the tubes between them effectively as a closed system.</p> <p>b) Should be compatible with all standard tubing with an external diameter ranging from 3.9-4.5 mm& internal diameter of 2.9-3.1 mm.</p> <p>c) Should be capable of generating a temperature of upto 320 degree Centi. To prevent any type of contamination.</p> <p>d) Should have in built sensor to monitor the temperature to ensure optimal quality and strength of weld.</p> <p>e) Should be capable of welding Wet - Wet / Wet - Dry / Dry - Dry tubes</p> <p>f) Total process time should be between 20 to 30 seconds.</p> <p>g) Requirement for tube length to be welding/docking should be as small as possible</p> <p>h) LED indicators to display the whole process with alarms.</p> <p>i) should ensure the complete sterility</p> <p>j) Should be attachable to leukocyte filters and should also be usable for plasma and platelet pooling</p> <p>k) There should be no particles or chemical residue created by welding process</p> <p>l) To be operational on 220 to 240 volts at 50 Hz. Single phase AC.</p> <p>m) Two thousand and five hundred consumables if any (eg. wafers) should be provided with instrument and cost of replacement consumables should be quoted and supply should be readily available with the vendor till the CMC period.</p> <p>4. Additional requirements</p>	

- a) Complete with comprehensive set of spare parts and a suitable capacity voltage stabilizer.
- b) Firm should supply one desk top computer (System should have i3 processor with latest operating system, HDD 1 TB, 4GB RAM, 17" screen) with laser printer with Split AC 1.5 Ton of repute make Free of cost with equipment. (Guarantee card with bills have to be submit with items to claim guarantee of these items)
- c) **Guarantee:** Three years on equipment from the date of installation.
- d) **CMC:** CMC will be given @ 5 % (of net rate - inclusive of Excise Duty & exclusive of VAT / CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of three years.
- e) The company should submit technical compliance sheet as per technical specifications mentioning the make & model of quoted item along with catalogue in the Technical bid.
- f) Installation will be done by supplier free of cost.
- g) The service engineer should be based in Rajasthan.
- h) Four Preventive Maintenance services annually during Guarantee and CMC period are essential.
- i) Demonstration of equipment is must for final technical approval.
- j) Consumable: Consumable kit. Rate of Consumable kit should be provided by bidder in BOQ.
- k) Final rate of consumable kit after negotiation will be freezed for two years

13. **Comprehensive Maintenance Contract Charges (per year/per machine)**

Rates (Rs.) Per Unit	First Year	Second Year	Third Year	Fourth Year	Fifth Year
Comprehensive Maintenance Agreement (CMA)	CMC: CMC will be given @ 5 % (of net rate - inclusive of Excise Duty & exclusive of VAT / CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after completion of Guarantee period of three years.				

~~Month~~


Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd.,
Jaipur

