

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

E-mail : edepmrmsc-ri@nic.in

CIN:U24232RJ2011SGC035067

Website:www.rmssc.nic.in

No. F-8(111) RMSC/EPM/M-1/NIB-190/RC.16-17/2017/130

Dated: 11/11/17

Rate Contract and Price Details

S.N.	Item	Detail												
1.	Name of Equipment / Item	pH Meter												
2.	Name & Address of Firm	M/s Chino Scientific Instruments Mfg. 3,15 Pancheel Nagar, Ajmer-305004 E-mail: mbranga@indiatimes.com , anirudhranga@live.com Tel: +91 145-2644858, 9414006868, 9460526867												
3.	Brand/ Model / Make	CSIM-304												
4.	Packing Unit	Each												
5.	R/C No. & Validity	R.C. No. 356 (Valid up to 31.01.2019)												
6.	Net Rate Per packing Unit (Rs.)	Rs. 6950.00												
7.	Taxes Consessional CST @ 2%/VAT @14.5% (Rs.)	VAT is Exempted												
8.	RMSC Surcharges @ 5%	Rs. 347.50												
9.	Total Price (Rs.)	Rs. 7297.50												
10.	Guarantee Period	Year :- One year, from the date of Installation of Equipment												
11.	CMC after Guarantee Period	CMC :- NIL												
12.	Technical specifications of Item pH Meter Type: Digital electronic pH Meter with combination pH electrodes 1. pH Range: 0.0 to 14.0 pH with 3 1/2 digital display 2. Mill volt Range: 0.0 to 1999 mv 3. Accuracy, repeatability and resolution: 0.01 pH 4. Calibration solution should be provided 5. Temperature compensation & pH standardization: 0° to 100°C(manual) 6. Electrical Power supply: 220 -240 volts at 50 Hz, single phase 7. Ambient Temperature: 0° to 40°C. 40°C at 95% RH 8. Equipment shall have suitable, rechargeable battery with a battery charger. 9. Type and details of battery being supplied shall be clearly indicated e.g. V, AH, number etc. 10. A suitable battery charger shall also be supplied so that charging batteries continue when the equipment on main. 11. Equipment shall be supplied in suitable case. 12. Technical literature: The firm shall positively submit printed illustrated literature/leaflet, indicating the model quoted by the firm. If quoted model is a modified version of their any standard product that also be indicated in the offer. 13. Guarantee: one year on equipment from the date of installation. 14. The company should submit technical compliance sheet as per technical specifications mentioning the make & model of quoted item along with catalogue in the Technical bid. 15. Installation will be done by supplier free of cost. 16. The service engineer should be based in Rajasthan. 17. Two Preventive Maintenance services annually during Guarantee and CMC period are essential. 18. Demonstration of item is must for final technical approval													
13.	Comprehensive Maintenance Contract Charges (per year/per machine)													
		<table border="1"> <thead> <tr> <th>Rates (Rs.) Per Unit</th> <th>First Year</th> <th>Second Year</th> <th>Third Year</th> <th>Fourth Year</th> <th>Fifth Year</th> </tr> </thead> <tbody> <tr> <td>Comprehensive Maintenance Agreement (CMA)</td> <td colspan="5">CMC: NIL</td> </tr> </tbody> </table>	Rates (Rs.) Per Unit	First Year	Second Year	Third Year	Fourth Year	Fifth Year	Comprehensive Maintenance Agreement (CMA)	CMC: NIL				
Rates (Rs.) Per Unit	First Year	Second Year	Third Year	Fourth Year	Fifth Year									
Comprehensive Maintenance Agreement (CMA)	CMC: NIL													

2.102
Executive Director (EPM)

Rajasthan Medical Services Corporation Ltd.,

Jaipur