## SPECIFICATIONS OF BATTERY OPERATED BONE DRILL

- 1. Autoclavable Drill with chuck and key.
- 2. It should be cannulated within 4 to 5.5 mm.
- 3. Forward and Reverse Movement.
- 4. The drill should have hand switch to control RPM.
- 5. Fully recharged battery should last for at least 3-4 hours (in continuous use).
- 6. Charging time should not be more than 04 Hours.
- 7. One spare battery should be provided.
- 8. System should be provided in case suitable for sterilization by formalin tablets.
- 9. RPM: more than 400 RPM.
- 10. Guarantee 03 years on complete system including battery.
- 11. CMC: CMC will be given @ 5% (of net rate- inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The rates of CMC should be provided for minimum five years after guarantee period of three years.
- 12. The equipment should be CE certified and ISO 13485. The bidder should be ISO 9001.
- 13. Consumable: Battery only.
- 14. Additional Key (Qty-01) should be provided.

Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd.,