

TECHNICAL SPECIFICATIONS
FRACTURE TABLE

1. It should have a stainless steel top grade 304.
2. It should be 7 feet long and 2 feet wide with four legs.
3. It should have proper reinforcement for the top.
4. It should be suitable for fracture treatment in orthopaedic operations.

ey

2 M

Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd
Jaipur