

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065
 CIN:U24232RJ2011SGC035067

E-mail : edepmrmisc-ri@nic.in
 Website: www.rmssc.health.rajasthan.gov.in

No. F-8(156)RMSC/EPM/M-3/16-17/NIB-191/ 3543

Date: 23/8/19

Order For Rate Change In Rate Contract Due To GST Provision

S.N.	Item	Detail
1.	Name of Equipment / Item	CTMT Machine
2.	Name & Address of Firm	M/s Medicaid Systems, 389, Industrial Area, Ph-II, Chandigarh-160 002 (India) Email: info@medicaid.co.in
3.	Brand/ Model / Make	MEDICAID'S CARDIVISION CV-USB
4.	Packing Unit	Each
5.	R/C No. & Validity	R.C. No. 359 (Valid up to 31.01.19)
6.	Net Rate Per packing Unit (Rs.)	Rs. 2,50,000.00 (In words Rs. Two Lacs Fifty Thousand Only)
7.	Taxes Concessional CST @ 2% (Rs.)	Rs. 5,000.00
8.	RMSC Surcharges @ 5%	Rs. 12,750.00
9.	Total Price (Rs.)	Rs. 2,67,750.00
10.	Basic New Rate (HSN Code 90189099)	
	(i) Existing Approved Net Rate in Rate Contract (In Rs.)	250000.00
	(ii) Less Excise Duty (In Rs.)	0.00
	(iii) Less any other amount intimated by the firm (In Rs.)	0.00
	(iv) New Basic Rate {(i)-(ii)-(iii)} (In Rs.)	250000.00
11.	Revised Rates for Supply to RMSC from Suppliers	
	(i) New Basic Rate (In Rs.)	250000.00
	(ii) IGST (In Rs.) 12%	30000.00
	(iii) CGST @% (In Rs.)	0.00
	(iv) SGST @% (In Rs.)	0.00
	(v) Approved Rate For Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (In Rs.)	280000.00



Revised Rate For Supply to Consignee From RMSC	
12.	
(i) New Basic Rate (In Rs.)	250000.00
(ii) RMSC Surcharge @ 5 % On (i) (In Rs.)	12500.00
(iii) CGST@ 6 % { On (i)+(ii)} (In Rs.)	15750.00
(iv) SGST@ 6% { On (i)+(ii)} (In Rs.)	15750.00
(v) Approved Rate For Consignee from RMSC {(i)+(ii)+(iii)+(iv)} (In Rs.)	294000.00

Executive Director (EPM)
RMSC, Jaipur

No. F-8(156)RMSC/EPM/M-3/16-17/NIB-191/3543

Date: 23/8/17

Copy forwarded to the following for information & necessary action:

1. P.S. to Mission Director NHM Jaipur.
2. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
3. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
4. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
5. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
6. All P.M.O.'s.
7. All CM&HO's.
8. Executive Director (Logistics/Finance)
9. AGM (IT), RMSC, Jaipur
10. All DPC's DDW Rajasthan.
11. M/s Medicaid Systems, 389, Industrial Area, Ph-II, Chandigarh-160 002 (India) Email: info@medicaid.co.in.
12. Guard file.

Executive Director (EPM)
RMSC, Jaipur

TECHNICAL SPECIFICATION OF CTMT MACHINE

1. The system should acquire 12 Lead ECG simultaneously & display 3x4, 4x6, 6x12, 12 lead data indicate significant changes.
2. Test Protocols: Bruce, Bruce Modified, Ellestad, Naughton and user defined program.
3. The system should have customizable summary report to determine the contents of report with data storage facility.
4. The system should be able to record resting ECG & Exercise ECG.
5. Should have alarm on HR, BP & ST level.
6. System should display at least these parameters :
 - a. 12 Lead raw ECG
 - b. 12 lead ST level & Slops
 - c. Enlarged QRS complex with maximum ST level
 - d. BP & HR
7. The system should have full facility to mark the ECG strip to enter comments at any stage.
8. The system should have facility to get system generated auto report.
9. The system should have facility to comparison between stages.
10. The system should have Grid at the background of training to measure the ST level manually.
11. The system should have following hardware & software details:
 - a. Data Acquisition : 12 Lead simultaneously
 - b. A/D : 12Bit or more
 - c. Sampling Rate : 300/Sec/channel or more
 - d. Input impedance : Should be more than 90M Ohms
 - e. CMRR : Should be more than 100 dB
 - f. Frequency response : 0.05Hz to 100Hz
 - g. Patient leakage : Should be less then 100 μ A
 - h. The system should have baseline correction.
 - i. Digital filters : 50Hz, 20Hz, 35Hz
 - j. Trade mill speed : 0.5 to 14km/Hr or higher range
 - k. Trade mill Motor : A.C./ D.C. Motor
 - l. Elevation : 0-21 or more
 - m. Communication : RS232/USB
 - n. Four tier heavy duty noise less trade mill with coasters having locking arrangement
 - o. Minimum Patient load 150 kg or more.
 - p. Emergency stop should be available on treadmill.
 - q. Walking area should be \geq 1300x400mm
12. The computer should have following configuration:
 - a. Processor : i5 or higher
 - b. Ram : 2GB or Higher
 - c. HDD : 500GB or Higher
 - d. DVD : Combo
 - e. Monitor : 17" or Higher
 - f. Keyboard : Multimedia
 - g. Printer : Laser printer of reputed brand
13. The system should he supplied with following items:
 - a. Good quality trolley system to mount computer & printer
 - b. Servo stabilizer of suitable rating & UPS for recording console
14. Manufacturer should have ISO 13485. Related documents should be submitted in technical bid.
15. The equipment should have CE certified. Related documents should be submitted in technical bid.
16. Equipment should conform to standards for electrical safety IEC-60601-1 General requirements and IEC-60601-2-25 Safety of Electrocardiograms. Related documents should be submitted in technical bid.

17. The firm should have service engineers based at Rajasthan.
18. Guarantee on accessories i.e. Lead set, (02nos.) should be minimum for 6 months.
19. Consumable: lead set (2nos.) & electrodes (100nos.) & other standard accessories: 10 rim reporting paper. System should be supplied with User/Operational Manual & service manual & other required literature in 2 copies at least for each.
20. Guarantee on complete system should be for three years.
21. CMC shall be awarded @5% of basic cost equipment + service tax (as applicable) with yearly escalation of 5% on last year's CMC price.
22. Computer system:- Make: Dell, Model: optiplex 3046 core i5-2.80 GHz.
23. Printer: Make: Samsung, Model: ML-1676P
24. Servo Stabilizer: Make Packard, Capacity: 3 KVA