



# Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065  
 CN:U24233RJ2011SGC035067

F-8(146)RMSC/EPM/M-3/16-17/NIB-183/

3484 Dated: 21-8-17

E-mail : edepmrmc-ri@nic.in  
 Website: www.rmhc.health.rajasthan.gov.in


## Order For Rate Change In Rate Contract Due To GST Provision

S.N.	Item	Detail
1.	Name of Equipment / Item	ICU Bed
2.	Name & Address of Firm	M/s United Surgical Industries, 445, F.I.E. Patparganj Industrial Area, Delhi-110092 (India) Email:-info@unisurg.com
3.	Brand/ Model / Make	United USI-3001H
4.	Packing Unit	Each
5.	R/C No. & Validity	R.C. No. 363 (Valid up to 28.2.2019)
6.	Net Rate Per packing Unit (Rs.)	85860.00 (In words Rs. Eighty Five Thousand Eight Hundred Sixty Only)
7.	Taxes Concessional CST @ 2% (Rs.)	Rs. 1717.20
8.	RMSC Surcharges @ 5%	Rs. 4378.86
9.	Total Price (Rs.)	Rs. 91956.06
10.		Basic New Rate (HSN Code 9402)
	(i) Existing Approved Net Rate in Rate Contract (In Rs.)	85860.00
	(ii) Less Excise Duty (In Rs.)	4860.00
	(iii) Less any other amount intimated by the firm (In Rs.)	0.00
	(iv) New Basic Rate {(i)-(ii)-(iii)} (In Rs.)	81000.00
11.		Revised Rates for Supply to RMSC from Suppliers
	(i) New Basic Rate (In Rs.)	81000.00
	(ii) ICGST (In Rs.) 18 %	14580.00
	(iii) CGST @ % (In Rs.)	0.00
	(iv) SGST @ % (In Rs.)	0.00
	(v) Approved Rate For Supplier To RMSC	95580.00
	{(i)+(ii)+(iii)+(iv)} (In Rs.)	



12.

Revised Rate For Supply to Consignee From RMSC	
(i) New Basic Rate (In Rs.)	81000.00
(ii) RMSC Surcharge @ 5 % On (i) (In Rs.)	4050.00
(iii) CGST@9 % { On (i)+(ii) } (In Rs.)	7654.50
(iv) SGST@ 9 % { On (i)+(ii) } (In Rs.)	7654.50
(v) Approved Rate For Consignee from RMSC {(i)+(ii)+(iii)+(iv)} (In Rs.)	100359.00

  
Executive Director (EPM)  
RMSC, Jaipur


F-8(146)RMSC/EPM/M-3/16-17/NIB-183/

3484

Dated: 21-8-17

Copy forwarded to the following for information & necessary action:

1. P.S. to Mission Director NHM Jaipur.
2. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
3. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
4. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
5. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
6. All P.M.O.'s.
7. All CM&HO's.
8. Executive Director (Logistics/Finance)
9. AGM (IT), RMSC, Jaipur
10. All DPC's DDW Rajasthan.
11. M/s United Surgical Industries, 445, F.I.E. Patparganj Industrial Area, Delhi-110092 (India) Email:-info@unisurg.com.
12. Guard file.

  
Executive Director (EPM)  
RMSC, Jaipur

### **TECHNICAL SPECIFICATION OF Intensive Care Bed (Motorized)**

1. Overall Size: 2200mm L x 1000mm W x 450mm to 750mm H  $\pm$ 100mm tolerance (without mattress).
2. Electrical System: The wired handset controls Hi-Lo Height adjustment, Trendelenburg/ Reverse trendelenburg, Back rest and Knee rest, Zero position for CPR.
3. The wired control panel for nurse shall be able to lock or operate: height, backrest, upper leg section, trendelenburg, reverse trendelenburg and zero position for CPR.
4. Battery Backup: The electrical system shall have a battery backup with built in battery charger, The backup should have 50 or more cycles.
5. Bed should have manual emergency CPR level on both sides.
6. The bed should have linear actuator based electrically operated height, knee break, backrest, trendelenburg/ reverse trendelenburg tech ( $\pm$ 12°) functions.
7. Construction: Top frame and base frame shall be mainly made from 50mm x 30mm x 1.6mm ERW (electric resistance welded) M.S. Rectangular tube.
8. Stainless steel wherever used shall conform to grade 304 (non magnetic). Support system linkages shall be of thick M.S. Flats/tubes.
9. Dust protective base cover shall be provided on control box.
10. The bed should have a provision to fix saline stand on all four corners of bed.
11. Detachable four-section top made from ABS material or epoxy powder coated perforated M.S. 1.6mm.
12. The bed should have safety side railings, collapsible split type shall be made from ABS. It should be easy to operate and have damping mechanism. Zero transfer gap facility shall be available.
13. Retractable back rest & knee rest.
14. Caster Wheels: The bed shall be provided with central locking 125mm diameter wheels and directional locking of wheels shall have thread guards; the castor body shall be made of non-rusting material. Corner Buffers: Four nos. rubber buffer shall be fitted to the four corners.
15. Oxygen Cylinder holder shall be provided.
16. Safe overall working load should be minimum 150kg.
17. Bed should have dual sided integral drainage bag rails with hooks.
18. The Bed should operate on 220-240V AC 50/60Hz.
19. Finish: All components should be thoroughly pre-treated chemically to remove rust and foreign matter like grease, oil, etc. by epoxy polyester powder with appropriate thickness paint film. This finish may exclude castor wheels any stainless steel parts, hardware, ebonite rubber and PVC parts used.
20. The product should be CE certified. Related documents shall be submitted along with technical bid.
21. The product IEC EN 60601-2-38/ 60601-2-52 international safety for electricity actuated hospital bed certified. Related documents should be enclosed in technical bid.
22. Supplied with:
  - a) Should be provided with foam mattress of 4 sections, 4" thick PU foam of 40 density covered with soft water proof material, bacteriostatic & pillow.
  - b) Stainless Steel Saline Rod -01 Number
23. **Guarantee:** should be three years from the date of successfully installation at consignee place.
24. **CMC:** CMC will be given @5% (of net rate-inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) and yearly escalation of 5% on last year's CMC price. The CMC will be awarded for five years (on yearly basis) after Guarantee period of three years.

### **Other Terms & Conditions:**

1. Manufacturer should be ISO- 13485 certified.
2. Final Technical approval after demonstration.
3. Installation will be done by supplier free of cost.
4. The Company should submit technical compliance sheet as per technical specifications mentioning the make and model of quoted item along with catalogue in technical bid.