

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

CIN:U24232RJ2011SGC035067

No. F-8()RMSC/EPM/M-2/2016-17/NIB-221/ 3419

E-mail : edepmrmisc-rj@nic.in

Website: www.rmhc.health.rajasthan.gov.in

Dated: 14/8/17

Order For Rate Change In Rate Contract Due To GST Provision

S.N.	Item	Detail
1.	Name of Equipment / Item	CTG Machine (Cardio tocography)-Fetal Monitor
2.	Name & Address of Firm	M/s Medical Care System, Manohar Villa, 3rd Floor, 280 behind Bhagat Singh Park, Adarsh Nagar, Jaipur-302004 E-Mail:-psk.medicalcare@gmail.com
3.	Brand/ Model / Make	Biocare (IC 90)
4.	Packing Unit	Each
5.	R/C No. & Validity	R.C. No. 376 (Valid up to 31.03.2019)
6.	Net Rate Per packing Unit (Rs.)	Rs 1,34,500.00 (In Words, Rs. One Lac Thirty Four Thousand Five Hundred Only) + VAT Exempted
7.	Concessional CST @ 2%	VAT Exempted
8.	RMSC Surcharges @ 5%	Rs. 6,725.00
9.	Total Approved Rate Including Concessional; CST and RMSC Surcharge	Rs. 1,41,225.00
10.	Basic New Rate (HSN Code(9018))	
	(i) Existing Approved Net Rate in Rate Contract (In Rs.)	
	(ii) Less Excise Duty (In Rs.)	134500.00
	(iii) Less any other amount intimated by the firm (In Rs.)	0.00
	(iv) New Basic Rate {(i)-(ii)-(iii)} (In Rs.)	0.00
		134500.00
11.	Revised Rates for Supply to RMSC from Suppliers	
	(i) New Basic Rate (In Rs.)	
	(ii) IGST (In Rs.)	134500.00
	(iii) CGST @ 6% (In Rs.)	0.00
	(iv) SGST @ 6% (In Rs.)	8070.00
	(v) Approved Rate For Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (In Rs.)	8070.00
		150640.00

[Handwritten Signature]

12. Revised Rate For Supply to Consignee From RMSC	
	134500.00
(i) New Basic Rate (In Rs.)	6725.00
(vi) RMSC Surcharge @ 5 % On (i) (In Rs.)	8473.50
(vii) CGST@ 6 % { On (i)+(ii)} (In Rs.)	8473.50
(viii) SGST@ 6 % { On (i)+(ii)} (In Rs.)	158172.00
(ix) Approved Rate For Consignee From RMSC {(i)+(ii)+(iii)+(iv)} (In Rs.)	

2/2
Executive Director (EPM)
RMSC, Jaipur

Dated: 16/10/17

No. F-8()RMSC/EPM/M-2/2016-17/NIB-221/ 3419

Copy forwarded to the following for information & necessary action:

1. P.S. to Mission Director NHM Jaipur.
2. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
3. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
4. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Scinces, Jaipur).
5. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
6. All P.M.O.'s.
7. All CM&HO's.
8. Executive Director (Logistics/Finance)
9. AGM (IT), RMSC, Jaipur
10. All DPC's DDW Rajasthan.
11. M/s Medical Care System, Manohar Villa, 3rd Floor, 280 behind Bhagat Singh Park, Adarsh Nagar, Jaipur-302004
12. Guard file.

2/2
Executive Director (EPM)
RMSC, Jaipur

Specifications of item CTG Machine (Cardio Tocography)-Fetal Monitor

1. Fetal monitor for recording and analyzing the Fetal Heart Rate (FHR) on beat-to-beat basis.
2. Toco and Maternally sensed fetal movements. both manually and automatically detected
3. Should have facility of twin monitoring.
4. Should have built-in printer with multiple printing speed and automatic paper feeding facility.
5. Fetal monitor should have inbuilt rechargeable battery of **atleast 60 min. backup.**
6. Display of FHR upto Twins FHR1 & FHR2 & TOCO on **10" or more display.**
7. Heart rate offset mode to visually distinguish between the heart rates of twins with ability to offset the secondary FHR by atleast 20 BPM.
8. Uterine contractions alarm. Alarm delay facility, so that alarm is available only if the alarming condition is persistent for preset time. User friendly alarms and patient data.
9. Actual FHR in BPM.
10. Blinking corresponding to each Beat.
11. Patient Name / ID, date and time, Memory Backup/Graphical or Tabular trend for minimum **24** hours with fast printing facility.
12. In-built / **Separate Acoustic** Stimulator with a separate marker on the graph for acoustic stimulators.
13. Ultrasound transducer should be multi crystal wide beam pulsed Doppler with frequency of 1MHz.
14. Fetal Heart Rate - measurements Range: 50-220 BPM.
15. External toco transducer which should be a sealed waterproof unit. Guard ring designed to reduce maternal respiration artifact. Measurement Range: 0-99 Units.
16. Event Market - Hand held, patient operated as well as front panel operated. Voltage 230V A.C. $\pm 10\%$, 50 Hz.
17. Product should be European CE / USFDA certified.
18. **Guarantee:** Three Years on equipment from the date of installation.
19. **CMC:** CMC will be given @ 5 % (of net rate- inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after guarantee period of three years.
20. Installation will be done by supplier free of cost.
21. List of spare parts and accessories to be supplied.
22. Company should have Service Network in Rajasthan.
23. The company should mention the make & model name/number of the quoted equipment and submit the technical brochure of the quoted model in the technical bid along with compliance sheet as per technical specifications.
24. Should be supplied with following accessories:-
 - a) FHR Probe - 2 Nos.
 - b) TOCO Probe - 1 No.
 - c) Event Marker - 1 No.
 - d) Fetal Simulator - 1 No.
 - e) Probe Belt - 3 Nos.
 - f) Ultrasound gel - 10 Nos.
 - g) Compatible Paper roll - 10 Nos.
 - h) Power Cable - 1 No.
 - i) Movable Trolley with cover - 1 No.
25. Rate of printing paper (Compatible with machine) should be provided by bidder in BOQ. Final rate of printing paper will be freezed after negotiation.
26. Final technical approval after demonstration on patient.