

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tikak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

CIN:U24232RJ2011ISGC035067

E-mail : edepnrmsc-f@nic.in

Website: www.health.rajasthan.gov.in

Dated: 29/01/19

Order for Rate Changes In Rate Contract Due To GST Provision

No. F-8() RMSC/EPMM-4/16-17/NIB-201/3567

S. No.	Item	Detail
1.	Name of Equipment / Item	MVA KTR (Syringe with accessories).
2.	Name & Address of Firm	M/s Global Medical Device, Kalasuruti Building, Ground Floor, S.NO. 46/2 Dattaprasad Housing Society, Kharadi Road, Chandanagar, Pune 14. E-mail: globalmedicaldevices@gmail.com Mobile No. 9890014348 Tel. 020-41208304 WOMENCARE GLOBAL
3.	Brand/ Model / Make	
4.	Packing Unit	Each
5.	R/C No. & Validity	(R.C. No 378 /Valid upto 30.04.2019)
6.	Net Rate Per packing Unit (Rs.)	1814.00 (In Rs One Thousand Eight Hundred Fourteen Only)
7.	Concessional CST @2%	Rs. 36.28
8.	RMSC Surcharge @5%	Rs. 92.51
9.	Total Approved Rate Including Concessional ; CST And RMSC Surcharge	Rs. 1942.80
10.	Basic New Rate (HSN Code 90183930)	
	(i) Existing Approved Net Rate in Rate Contract (In Rs.)	1814.00
	(ii) Less : Excise Duty (In Rs.)	0.00
	(iii) Less: any other amount intimated by the firm (In Rs.)	0.00
	(iv) New Basic Rate {(i)-(ii)-(iii)} (In Rs.)	1814.00

Revised Rates for Supply to RMSC from Suppliers

11.		Revised Rates for Supply to RMSC from Suppliers	
(i)	New Basic Rate (In Rs.)	1814.00	
(ii)	IGST @ 12 % (In Rs.)	217.68	
(iii)	CGST (In Rs.)	0.00	
(iv)	SGST (In Rs.)	0.00	
(v)	Approved Rate for Consignee From RMSC {(i)+(ii)+(iii)+(iv)} (In Rs.)	2031.68	
Revised Rate for Supply To Consignee From RMSC			
12.			
(i)	New Basic Rate (In Rs.)	1814.00	
(ii)	RMSC Surcharge @ 5% on (In Rs.)	90.70	
(iii)	CGST @ 6% {On (i)+(ii)} (In Rs.)	114.28	
(iv)	SGST @ 6% {On (i)+(ii)} (In Rs.)	114.28	
(v)	Approved Rate for Consignee From RMSC {(i)+(ii)+(iii)+(iv)} (In Rs.)	2133.26	

Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd,
Jaipur

Dated 24/8/17

No. F-8() RMSC/EPM/M-4/16-17/NIB-201/ 3567
Copy forwarded to the following for information & necessary action:

1. P.S. to Mission Director NHM Jaipur.
2. Director: (PH)/(Aids)/(FW)/(RCH)/(EC)/(ESI).
3. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
4. Superintendent Japunya Hospital Jaipur (Associated with RJHS College of Medical Sciences, Jaipur).
5. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Jaipur.
6. All P.M.O.'s
7. All CM&HO's
8. Executive Director (Logistics/Finance)
9. AGM (TD), RMSC, Jaipur
10. All DPCs DDW Rajasthan.
11. M/s Global Medical Device, Kalasuruti Building, Ground Floor, S.NO. 46/2 Dattaprasad Housing Society, Kharadi Road, Chandannagar, Pune
14. E-mail: globalmedicaldevices@gmail.com
12. Guard file.

Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd,
Jaipur

Technical specifications of MVA KIT (Syringe with accessories)

1. Syringe Type: Double valve manual vacuum aspiration syringe
2. Material: Plastic
3. Capacity: 50-60 cc.
4. Pressure: 24-26 inches of Mercury
5. Plunger: Double arm plunger with "O" Ring Collars stop along with retaining clips.
6. Valve: Made of Silicon double pinch valve with a liner.
7. Cannula: Malleable plastic, 4, 5, 6, 7, 8, 9, 10 and 12 mm (total = 8) in size, in 5 different colours.
8. Compatible with the syringe without an adapter.
9. Additional three (3) "O" rings & Silicon oil for lubrication, 3 (three) extra collar stop to be supplied with each kit.
10. Kit to be supplied in transparent plastic box which is also autoclavable.
11. Sterilization: All parts of the syringe should be cold sterilizable with Glutaraldehyde & 0.5% Chlorine, should be also autoclavable at 121°C
12. Reusability: The kit should be reusable at least for 25 times under defined sterilization conditions. A certificate in this regard should be furnished from a Government approved lab.
13. Shelf life: Minimum 2 yrs. from the date of delivery.
14. Manufacturer should be ISO 13485 certified.