

Rajasthan Medical Services Corporation Limited, Jaipur
 Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph.No. 0141-2223887, Fax : 0141-2228065
 CIN:U24232RJ2011SGC035067

E-mail : edepmrmisc-ri@nic.in
 Website: www.rmhc.health.rjasthan.gov.in


No. F-8 (27) RMSC/EPM/M-5/2016-17/NIB-177/

Dated:

Order for Rate Changes In Rate Contract Due To GST Provision

S. No.	Item	Detail
1.	Name of Equipment / Item	Orthopaedic OT Table (Electric)
2.	Name & Address of Firm	M/s Surgident (India), A-119, Mangolpuri Industrial Area, Phase-II, New Delhi-110083 Email: surgidentindia@yahoo.in
3.	Brand/ Model / Make	Surgident / Electromac 1
4.	Packing Unit	One
5.	R/C No. & Validity	R.C. No- 380 & Valid upto 30.04.2019)
6.	Net Rate Per packing Unit (Rs.)	Rs. 2,75,600.00 (In words Rs. Two Lakh Seventy Five Thousand Six Hundred Only)
7.	Concessional CST @2%	Rs. 5,512.00
8.	RMSC Surcharge @5%	Rs. 14,056.00
9.	Total Approved Rate Including Concessional ; CST And RMSC Surcharge	Rs. 2,95,168.00
10.	Basic New Rate (HSN Code 94021010)	
	(i) Existing Approved Net Rate in Rate Contract (In Rs.)	275600.00
	(ii) <u>Less</u> : Excise Duty (In Rs.)	15600.00
	(iii) <u>Less</u> : Input Tax amount intimated by the firm 4% (In Rs.)	0.00
	(iv) New Basic Rate {(i)-(ii)-(iii)} (In Rs.)	260000.00

11. Revised Rates for Supply to RMSC from Suppliers		
(i)	New Basic Rate (In Rs.)	260000.00
(ii)	IGST @ 18% (In Rs.)	46800.00
(iii)	CGST 0% (In Rs.)	0.00
(iv)	SGST 0% (In Rs.)	0.00
(v)	Approved Rate for Suppliers To RMSC {(i)+(ii)+(iii)+(iv)} (In Rs.)	306800.00
12. Revised Rate for Supply To Consignee From RMSC		
(i)	New Basic Rate (In Rs.)	260000.00
(ii)	RMSC Surcharge @ 5% on (i) (In Rs.)	13000.00
(iii)	CGST @ 9% {On (i)+(ii)} (In Rs.)	24570.00
(iv)	SGST @ 9% {On (i)+(ii)} (In Rs.)	24570.00
(v)	Approved Rate for Consignee From RMSC {(i)+(ii)+(iii)+(iv)} (In Rs.)	322140.00


Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd.,
Jaipur
 Dated: 9/10/17

No. F-8 (27) RMSC/EPM/M-5/2016-17/NIB-1771 5013

Copy forwarded to the following for information & necessary action:

1. P.S. to Mission Director NHM Jaipur.
2. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
3. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
4. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Sciences, Jaipur).
5. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
6. All P.M.O.'s.
7. All CM&HO's.
8. Executive Director (Logistics/Finance)
9. AGM (IT), RMSC, Jaipur
10. All DPC's DDW Rajasthan.
11. M/s Surgident (India), A-119, Mangolpuri Industrial Area, Phase-II, New Delhi-110083
12. Guard File.


Executive Director (EPM)
Rajasthan Medical Services Corporation Ltd.,

Technical Specification of OT Table (Electric)

Measurements (Approx) : without mattress

Height: 700-1000mm \pm 50mm (without mattress)

Width: 500 mm or more

Length: 2000 mm or more

Maximum patient weight capacity 150 kg or more

Cranial and caudal traversing 200 mm or better

Side tilt: \pm 20 degrees,

Back section adjustment: -40 degrees to 80 degrees,

Foot section adjustment: -90 to degree, detachable,

Trendelenburg: 20 degree \pm 5 degree, Anti trendelenburg: 20 degree \pm 5 degree

Head section adjustment: \pm 45 degree, detachable

The maximum distance from perineal post to foot attachment should be more than 1200mm

The operation table should have the following features:-

Four/Five section table top with divided foot section, Table top should permit X-Ray penetration and fluoroscopy. The frame should be designed in such a way for minimum obstruction for C-Arm imaging. Mattress should be radiolucent and suitable for fluoroscopy.

Should have a handset for various functions. There should be inbuilt standby control enabling full use of table in case of handset failure. All table positioning, i.e., height, back section, lateral tilt, trendelenburg, and anti-trendelenburg, should be operated electrically through handset and standby control. The table top can be moved cranially and caudally on its base.

Electric actuators should be good quality European/Japan make.

The casing on the frame and made of hygienic stainless steel. Table should have mobile base with lockable castors with electric/mechanical floor locking mechanism, The centre supporting column should be eccentrically placed on the base.

The table should have a battery back-up of at least 30 minutes.

The table should be provided with the following standard accessories:

- a. Anaesthesia screen double ended with clamp: 1 No.
- b. Body strap: 1no.
- c. Knee crutch with clamp: -1pair
- d. Shoulder support - 1 pair
- e. Side support – 1 pair
- f. Padded arm rest with straps – one pair
- g. Polyurethane mattress 50 mm thick with Volero attachment

O.T. Table should be provided with the following orthopaedic attachments

- a. Fracture attachment (attached to base, non hanging).
- b. Perennial post with pad – adult and pediatric – one each.

- c. Sacral rest wedge shaped.
- d. Hip rest.
- e. Counter traction device for femur and tibia nailing.
- f. L-shaped Traction extension
- g. Traction bow.
- h. Universal Leg holder – adult and pediatric – one each.
- i. Multi-purpose Arm Board with cushion and clamp.

Guarantee: 03 year from the date of successful installation.

CMC: CMC will be given @ 4 % (of net rate-inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) and yearly escalation of 5 % on last year's CMC price. The CMC will be awarded for five years on yearly basis after Guarantee period of three years.

The equipment should be CE certified and ISO 13485.