

Rajasthan Medical Services Corporation Limited, Jaipur

Gandhi Block, Swasthya Bhawan, Tilak Marg, C-Scheme, Jaipur - 302005

Ph. No. 0141-2223887, Fax : 0141-2228065

CIN:U24232RJ2011SGC035067

E-mail : edepmrmc-ri@nic.in

Website: www.rmhc.health.rajasthan.gov.in

F No. F-8(211) RMSC/EPM/M-3/16-17/NIB 215/

3475 Date:- 21-8-17

Order For Rate Change In Rate Contract Due To GST Provision

S.N.	Item	Detail
1.	Name of Equipment / Item	3 Channel ECG Machine
2.	Name & Address of Firm	M/s Recorders & Medicare Systems (P) Ltd., 196, Industrial Area, Phase 1, Panchkula (Haryana)-134113(India):- helpdesk@rmsindia.com
3.	Brand/ Model / Make	"RMS" Vesta 301i
4.	Packing Unit	Each
5.	R/C No. & Validity	R.C. No. 407 (Valid up to 31.7.2019)
6.	Net Rate Per packing Unit (Rs.)	Rs. 27250.00
7.	Taxes Concessional CST @ 2 % (Rs.)	Rs. 545.00
8.	RMSC Surcharges @ 5%	Rs. 1389.75
9.	Total Price (Rs.)	Rs. 29184.75
10.	Basic New Rate (HSN Code 90181100)	
	(i) Existing Approved Net Rate in Rate Contract (In Rs.)	27250.00
	(ii) Less Excise Duty (In Rs.)	1542.45
	(iii) Less any other amount intimated by the firm (In Rs.)	0.00
	(iv) New Basic Rate {(i)-(ii)-(iii)} (In Rs.)	25707.55
11.	Revised Rates for Supply to RMSC from Suppliers	
	(i) New Basic Rate (In Rs.)	25707.55
	(ii) IGST (In Rs.) 12%	3084.91
	(iii) CGST @% (In Rs.)	0.00
	(iv) SGST @% (In Rs.)	0.00



	(v) Approved Rate For Supplier To RMSC {(i)+(ii)+(iii)+(iv)} (In Rs.)	28792.46
12.	Revised Rate For Supply to Consignee From RMSC	
	(i) New Basic Rate (In Rs.)	25707.55
	(ii) RMSC Surcharge @ 5 % On (i) (In Rs.)	1285.38
	(iii) CGST@6 % { On (i)+(ii)} (In Rs.)	1619.575
	(iv) SGST@6 % { On (i)+(ii)} (In Rs.)	1619.575
	(v) Approved Rate For Consignee from RMSC {(i)+(ii)+(iii)+(iv)} (In Rs.)	30232.08

Executive Director (EPM)
RMSC, Jaipur

F No. F-8(211) RMSC/EPM/M-3/16-17/NIB 215/

3479 Date:- 21-8-17

Copy forwarded to the following for information & necessary action:

1. P.S. to Mission Director NHM Jaipur.
2. Director, (PH)/(Aids)/(FW)/(RCH)/(IEC)/(ESI).
3. Principal, Medical College, Ajmer, Bikaner, Jaipur, Jodhpur, Kota, Jhalawar & Udaipur.
4. Superintendent Japuriya Hospital Jaipur (Associated with RUHS College of Medical Scinces, Jaipur).
5. Director, Medical & Health Services (ESI), Mobile Surgical Unit, Raj. Jaipur.
6. All P.M.O.'s.
7. All CM&HO's.
8. Executive Director (Logistics/Finance)
9. AGM (IT), RMSC, Jaipur
10. All DPC's DDW Rajasthan.
11. M/s Recorders & Medicare Systems (P) Ltd., 196, Industrial Area, Phase 1, Panchkula (Haryana)-134113(India):- helpdesk@rmsindia.com.
12. Guard file.

Executive Director (EPM)
RMSC, Jaipur

Technical Specification for 3 Channel ECG Machine

1. 3 channel ECG Machine with LCD display, Size 4" or more.
2. Automatic, Manual Recording Modes.
3. Soft touch key Board.
4. Paper size: 60 mm and more.
5. Mains and Rechargeable battery operated.
6. Inbuilt Rechargeable battery for at least 2 hours.
7. Portable, lightweight < 3 kg.
8. Measurement facility with interpretation.
9. Memory for 25 ECG or more.
10. Automatic Lead Marking, Thermal printing.
11. Recording Sensitivity: 2.5, 5, 10, 20 mm/mV.
12. Defibrillation Protection: Internal
13. Frequency response: 0.05-150 Hz.
14. Filters: Mains and Muscle interference.
15. Safety standards : Class II, type CF.
16. CE (European) or USFDA approved.
17. Guarantee 3 years on equipment from the date of installation & minimum 06 months on Consumables.
18. CMC: CMC shall be given @ 5% of net rate (inclusive of Excise Duty & exclusive of VAT/CST etc.) plus service tax (as applicable) and yearly escalation of 5% on last year's CMC price. The CMC may be awarded for five years (on yearly basis) after Guarantee period of three years.
19. Equipment should be supplied with compatible standard paper rolls (quantity 10 nos.), additional Clamp Electrodes (1 set), Chest Electrodes (1 set), ECG Jelly (250 ml X 2 bottles), ECG Cables (1 set) , Power Cable (1 set) Earth Cable-1, Guarantee Card & Operating Manual.
20. Consumable: Battery, Lead Set, Chest Electrodes, Limb Electrodes, ECG Jelly and ECG Paper, Rates of Battery, Lead Set, Chest Electrodes and Limb Electrodes should be provided by bidder in BOQ (BF-IV).
21. The rates of consumables and spare parts should be provided and that will be freezed for minimum three years.
22. The machine should be supplied with a good quality carry bag to carry ECG Machine.
23. Detail of Service Network of Firm in Rajasthan.
24. The Company should mention the model name/number and submit the technical brochure of the quoted model in the Technical Bid.